



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB123597

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	01/09/2023	17:10
2	CAL2	1	Water	NA	NA	20.8	7.00	01/09/2023	17:11
3	CAL3	1	Water	NA	NA	20.8	10.03	01/09/2023	17:15
4	ICV	1	Water	NA	NA	20.7	6.98	01/09/2023	17:20
5	CCV1	1	Water	NA	NA	20.7	2.01	01/09/2023	17:22
6	O1076-01	1	Solid	20.02	20	20.2	8.75	01/09/2023	17:25
7	O1076-01DUP	1	Solid	20.03	20	20.3	8.76	01/09/2023	17:26
8	CCV2	1	Water	NA	NA	20.7	12.03	01/09/2023	17:30

VB 123597

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O1082 S WorkList ID : 166436 Department : Wet-Chemistry Date : 01-09-2023 16:45:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/09/2023	Solid	O1076-01	pH					01/09/2023	9045D

Date/Time 01.09.23 16:50
Raw Sample Received by: 20 (COC)
Raw Sample Relinquished by: AL SPZ

Date/Time 01.09.23 18:00
Raw Sample Received by: AL SPZ
Raw Sample Relinquished by: 70 (COC)