



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123600

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	01/10/2023	12:00
2	O1077-01		1	12	100.00	>212.0	>212.0	01/10/2023	12:30
3	O1077-01DUP		1	12	100.00	>212.0	>212.0	01/10/2023	13:00
4	O1079-01		1	14	100.00	>212.0	>212.0	01/10/2023	13:30
5	O1079-02		1	13	100.00	>212.0	>212.0	01/10/2023	14:00
6	O1081-04		1	12	100.00	>212.0	>212.0	01/10/2023	14:30
7	O1081-05		1	14	100.00	>212.0	>212.0	01/10/2023	15:00
8	O1081-06		1	14	100.00	>212.0	>212.0	01/10/2023	15:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16123600

WorkList Name : fp1-10

WorkList ID : 166439

Department : Wet-Chemistry

Date : 01-10-2023 08:30:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/12/2023	Water	O1077-01	Flash Point	Cool 4 deg C	PSEG03	N32	BUR-1550	01/09/2023	1010B
01/10/2023	Water	O1079-01	Flash Point	Cool 4 deg C	SOUN03	N31	SAMPLE-1-EFFLUENT	01/09/2023	1010B
01/10/2023	Water	O1079-02	Flash Point	Cool 4 deg C	SOUN03	N31	SAMPLE-2-INFLUENT	01/09/2023	1010B
01/12/2023	Water	O1081-04	Flash Point	Cool 4 deg C	PSEG03	N32	599	01/09/2023	1010B
01/12/2023	Water	O1081-05	Flash Point	Cool 4 deg C	PSEG03	N32	601	01/09/2023	1010B
01/12/2023	Water	O1081-06	Flash Point	Cool 4 deg C	PSEG03	N32	31944	01/09/2023	1010B

Date/Time 01/10/2023 11:50
 Raw Sample Received by: RPN CWC
 Raw Sample Relinquished by: RJ CEA-1046

Date/Time 01/10/2023 15:35
 Raw Sample Received by: RJ CEA-1046
 Raw Sample Relinquished by: RPN CWC