



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123601

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	01/10/2023	10:30
2	O1082-01		1	12	100.00	>212.0	>212.0	01/10/2023	11:00
3	O1082-01DUP		1	14	100.00	>212.0	>212.0	01/10/2023	11:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fps-1-10

WorkList ID : 166453

Department : Wet-Chemistry

Date : 01-10-2023 09:27:42

LB123601

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/12/2023	Solid	O1082-01	Flash Point	Cool 4 deg C	PSEG03	N33	585-OIL	01/09/2023	1010B

Date/Time 01/10/2023 9:25
 Raw Sample Received by: RM (WC)
 Raw Sample Relinquished by: RJ C EST-166)

Date/Time 01/10/2023 11:50
 Raw Sample Received by: RJ C EST-166)
 Raw Sample Relinquished by: RM (WC)