



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: Non-Polar Material
Run Number: LB123602
Analysis Date: 01/10/2023
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: Iwona
Extraction Date: 01/10/2023
Extraction IN Time: 08:30
Extraction OUT Time: 09:00
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB123602BL	LB123602BL	WATER	1.3	1000	100	2.5694	2.5694	3.01	2.5695	2.5695	0.0001	0.1
2	LB123602BS	LB123602BS	WATER	1.3	1000	100	2.9741	2.9741	3.02	2.9907	2.9907	0.0166	16.6
3	LB123602BSD	LB123602BSD	WATER	1.3	1000	100	3.0152	3.0152	3.02	3.0322	3.0322	0.0170	17
4	O1079-01	SAMPLE-1-EFFLUENT	WATER	1.3	1000	100	2.9896	2.9896	3.01	2.9901	2.9901	0.0005	0.5
5	O1079-02	SAMPLE-2-INFLUENT	WATER	1.3	1000	100	2.9887	2.9887	3.03	2.9899	2.9899	0.0012	1.2



QC Batch# LB123602 Test: Non-Polar Material Analysis Date: 01/10/2023

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2969
pH Paper 0-14	M4909
Sodium Sulfate	EP2279
1:1 HCL	WP10035
Silica Gel	W2839
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP100036
LCSWD	5.00 ML	WP100037
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

Before Analysis

0.0020 gram Balance: 0.0018 (0.0018-0.0022) In OVEN TEMP1 : 70 °C Dessicator Time In1 : 10:01

1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time1: 09:25

Bal Check Time: 08:40 Out OVEN TEMP1: 70 °C Dessicator Time Out1: 10:30

Out Time1: 10:00

After Analysis

0.0020 gram Balance: 0.0021 (0.0018-0.0022) In OVEN TEMP2 : 71 °C Dessicator Time In2 : 11:31

1.0000 gram Balance: 1.0005 (0.9950-1.0050) In Time2: 11:00

Bal Check Time: 12:10 Out OVEN TEMP2: 71 °C Dessicator Time Out2: 12:00

Out Time2: 11:30

18 AB 2306270
18123602

WORKLIST(Hardcopy Internal Chain)

WorkList Name : NON POPLAR O1069 WorkList ID : 166437 Department : Wet-Chemistry Date : 01-10-2023 08:17:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/10/2023	Water	O1079-01	Non-Polar Material	☒	Conc H2SO4 to	SOUN03	N31	SAMPLE-1-EFFLUENT	01/09/2023 1664A
01/10/2023	Water	O1079-02	Non-Polar Material	☒	Conc H2SO4 to	SOUN03	N31	SAMPLE-2-INFLUENT	01/09/2023 1664A

Date/Time 01.10.23 08:20
Raw Sample Received by: 7060C1
Raw Sample Relinquished by: CR SM

Date/Time 01.10.23 11:00
Raw Sample Received by: CR SM
Raw Sample Relinquished by: 7060C1