

LB123607


7471B

INSTRUMENT ID: CY4

Linear

 μ Abs.:

56087



A= 0.0000e+000

B= 1.7834e-004

C= 1.7551e-002

Rho= 0.9999831

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.026	0.026	47	0.000	47					0/0
0.2	0.200	0.204	0.004	1045	0.0 %	1045					1
2.5	2.500	2.485	-0.015	13835	0.0 %	13835					2
5.0	5.000	4.962	-0.038	27726	0.0 %	27726					-1
7.5	7.500	7.503	0.003	41972	0.0 %	41972					-1
10.0	10.000	10.020	0.020	56087	0.0 %	56087					0

LB123607 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 10 Jan 2023 12:12:05

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	47	-	0.0000	7471B	PPB	10 Jan 2023 12:24:23	S	Std
0.2 - 1	50.2	1045	-	0.2000	7471B	PPB	10 Jan 2023 12:26:41	S	Std
2.5 - 1	52.5	13835	-	2.5000	7471B	PPB	10 Jan 2023 12:28:58	S	Std
5.0 - 1	55	27726	-	5.0000	7471B	PPB	10 Jan 2023 12:31:15	S	Std
7.5 - 1	57.5	41972	-	7.5000	7471B	PPB	10 Jan 2023 12:33:31	S	Std
10.0 - 1	511	56087	-	10.0000	7471B	PPB	10 Jan 2023 12:35:47	S	Std
ICV97 - 1	ICV97	21313	3.8185	-	7471B	PPB	10 Jan 2023 12:38:31	U	SMPL
ICB97 - 1	ICB97	0	0.0176	-	7471B	PPB	10 Jan 2023 12:40:45	U	SMPL
CCV30 - 1	CCV30	27762	4.9686	-	7471B	PPB	10 Jan 2023 12:43:01	U	SMPL
CCB30 - 1	CCB30	-14	0.0151	-	7471B	PPB	10 Jan 2023 12:45:16	U	SMPL
CRA - 1	CRA	1087	0.2114	-	7471B	PPB	10 Jan 2023 12:47:31	U	SMPL
HighStd - 1	HighStd	56332	10.0638	-	7471B	PPB	10 Jan 2023 12:49:46	U	SMPL
ChkStd - 1	ChkStd	39065	6.9844	-	7471B	PPB	10 Jan 2023 12:52:01	U	SMPL
PB150151BL - 1	PB150151BL	-62	0.0065	-	7471B	PPB	10 Jan 2023 12:54:17	U	SMPL
PB150151BS - 1	PB150151BS	20578	3.6874	-	7471B	PPB	10 Jan 2023 12:56:33	U	SMPL
O1051-01 - 1	72-11933	2185	0.4072	-	7471B	PPB	10 Jan 2023 12:58:50	U	SMPL
O1051-01DUP - 1	72-11933DUP	2155	0.4019	-	7471B	PPB	10 Jan 2023 13:01:07	U	SMPL
O1051-01MS - 1	72-11933MS	21132	3.7862	-	7471B	PPB	10 Jan 2023 13:03:24	U	SMPL
O1051-01MSD - 1	72-11933MSD	20557	3.6837	-	7471B	PPB	10 Jan 2023 13:05:41	U	SMPL
O1057-01 - 1	COMPOSITE	6472	1.1718	-	7471B	PPB	10 Jan 2023 13:07:56	U	SMPL
CCV31 - 1	CCV31	27610	4.9415	-	7471B	PPB	10 Jan 2023 13:10:11	U	SMPL
CCB31 - 1	CCB31	-44	0.0097	-	7471B	PPB	10 Jan 2023 13:12:27	U	SMPL
O1061-05 - 1	SLUDGE-COMP	5216	0.9478	-	7471B	PPB	10 Jan 2023 13:14:42	U	SMPL
O1061-08 - 1	1623-1624	19	0.0209	-	7471B	PPB	10 Jan 2023 13:16:57	U	SMPL
O1061-09 - 1	12123	155	0.0452	-	7471B	PPB	10 Jan 2023 13:19:13	U	SMPL
O1064-01 - 1	S4-01	1943	0.3641	-	7471B	PPB	10 Jan 2023 13:21:29	U	SMPL
O1064-02 - 1	S4-02	1116	0.2166	-	7471B	PPB	10 Jan 2023 13:23:44	U	SMPL
O1064-03 - 1	S4-03	1341	0.2567	-	7471B	PPB	10 Jan 2023 13:26:01	U	SMPL
O1064-04 - 1	S-13-01	1784	0.3357	-	7471B	PPB	10 Jan 2023 13:28:18	U	SMPL
O1064-05 - 1	S-13-02	2291	0.4261	-	7471B	PPB	10 Jan 2023 13:30:35	U	SMPL
O1064-06 - 1	S-14-01	2475	0.4589	-	7471B	PPB	10 Jan 2023 13:32:52	U	SMPL
O1064-07 - 1	S-14-02	772	0.1552	-	7471B	PPB	10 Jan 2023 13:35:08	U	SMPL
CCV32 - 1	CCV32	27670	4.9522	-	7471B	PPB	10 Jan 2023 13:37:23	U	SMPL
CCB32 - 1	CCB32	-44	0.0097	-	7471B	PPB	10 Jan 2023 13:39:39	U	SMPL
O1064-08 - 1	S-14-03	1691	0.3191	-	7471B	PPB	10 Jan 2023 13:41:55	U	SMPL
O1065-01 - 1	VNJ-210	5899	1.0696	-	7471B	PPB	10 Jan 2023 13:44:10	U	SMPL
O1067-01 - 1	31920	200	0.0532	-	7471B	PPB	10 Jan 2023 13:46:26	U	SMPL
O1067-02 - 1	7822	3398	0.6236	-	7471B	PPB	10 Jan 2023 13:48:41	U	SMPL
O1069-01 - 1	TP-7	2784	0.5141	-	7471B	PPB	10 Jan 2023 13:50:57	U	SMPL
O1081-01 - 1	31963	6668	1.2067	-	7471B	PPB	10 Jan 2023 13:53:14	U	SMPL
O1081-02 - 1	31964	9021	1.6264	-	7471B	PPB	10 Jan 2023 13:55:30	U	SMPL
O1081-03 - 1	31965	879	0.1743	-	7471B	PPB	10 Jan 2023 13:57:47	U	SMPL
O1051-01LX5 - 1		419	0.0923	-	7471B	PPB	10 Jan 2023 14:00:04	U	SMPL
O1051-01A - 1		23217	4.1581	-	7471B	PPB	10 Jan 2023 14:02:20	U	SMPL
CCV33 - 1	CCV33	27841	4.9827	-	7471B	PPB	10 Jan 2023 14:06:46	U	SMPL
CCB33 - 1	CCB33	-36	0.0111	-	7471B	PPB	10 Jan 2023 14:09:02	U	SMPL