



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB123610

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	01/10/2023	13:15
2	CAL2	1	Water	NA	NA	20.8	7.00	01/10/2023	13:16
3	CAL3	1	Water	NA	NA	20.8	10.03	01/10/2023	13:18
4	ICV	1	Water	NA	NA	20.7	6.98	01/10/2023	13:20
5	CCV1	1	Water	NA	NA	20.8	2.01	01/10/2023	13:22
6	O1094-01	1	Solid	20.02	20	21.0	7.08	01/10/2023	13:25
7	O1094-01DUP	1	Solid	20.03	20	21.1	7.09	01/10/2023	13:26
8	O1094-03	1	Solid	20.05	20	21.3	8.19	01/10/2023	13:30
9	CCV2	1	Water	NA	NA	20.8	12.03	01/10/2023	13:33

VB 123610

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O1094 WorkList ID : 166468 Department : Wet-Chemistry Date : 01-10-2023 12:55:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/10/2023	Solid	O1094-01	pH	A	Cool 4 deg C	DASK01	N41	STOCKPILE-A	01/10/2023 9045D
01/10/2023	Solid	O1094-03	pH	A	Cool 4 deg C	DASK01	N41	STOCKPILE-B	01/10/2023 9045D

Date/Time 01.10.23 13:10
Raw Sample Received by: JD (wec)
Raw Sample Relinquished by: JD (wec)

Date/Time 01.10.23 15:00
Raw Sample Received by: JD (wec)
Raw Sample Relinquished by: JD (wec)