

LB123614


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

46954



A= 0.0000e+000

B= 2.1267e-004

C= 1.2790e-002

Rho= 0.9999641

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.023	0.023	46	0.000	46					9
0.2	0.200	0.218	0.018	965	0.0 %	965					3
2.5	2.500	2.432	-0.068	11376	0.0 %	11376					0
5.0	5.000	5.016	0.016	23527	0.0 %	23527					0
7.5	7.500	7.513	0.013	35266	0.0 %	35266					0
10.0	10.000	9.998	-0.002	46954	0.0 %	46954					0

LB123614 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 10 Jan 2023 14:30:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	46	-	0.0000	7470A	PPB	10 Jan 2023 14:39:01	S	Std
0.2 - 1	50.2	965	-	0.2000	7470A	PPB	10 Jan 2023 14:41:18	S	Std
2.5 - 1	52.5	11376	-	2.5000	7470A	PPB	10 Jan 2023 14:43:35	S	Std
5.0 - 1	55	23527	-	5.0000	7470A	PPB	10 Jan 2023 14:45:52	S	Std
7.5 - 1	57.5	35266	-	7.5000	7470A	PPB	10 Jan 2023 14:48:09	S	Std
10.0 - 1	510	46954	-	10.0000	7470A	PPB	10 Jan 2023 14:50:25	S	Std
ICV98 - 1	ICV98	17602	3.7562	-	7470A	PPB	10 Jan 2023 14:53:09	U	SMPL
ICB98 - 1	ICB98	54	0.0243	-	7470A	PPB	10 Jan 2023 14:57:58	U	SMPL
CCV34 - 1	CCV34	25375	5.4092	-	7470A	PPB	10 Jan 2023 15:00:12	U	SMPL
CCB34 - 1	CCB34	-34	0.0056	-	7470A	PPB	10 Jan 2023 15:02:28	U	SMPL
CRA - 1	CRA	944	0.2135	-	7470A	PPB	10 Jan 2023 15:04:43	U	SMPL
HighStd - 1	HighStd	44118	9.3953	-	7470A	PPB	10 Jan 2023 15:06:58	U	SMPL
ChkStd - 1	ChkStd	32530	6.9309	-	7470A	PPB	10 Jan 2023 15:09:14	U	SMPL
PB150152BL - 1	PBW	-33	0.0058	-	7470A	PPB	10 Jan 2023 15:11:30	U	SMPL
PB150152BS - 1	LCSW	18389	3.9235	-	7470A	PPB	10 Jan 2023 15:13:46	U	SMPL
O1053-02 - 1	TP-201	10	0.0149	-	7470A	PPB	10 Jan 2023 15:16:02	U	SMPL
O1054-01 - 1	COMP-1	39	0.0211	-	7470A	PPB	10 Jan 2023 15:18:19	U	SMPL
O1054-02 - 1	COMP-2	20	0.0170	-	7470A	PPB	10 Jan 2023 15:20:36	U	SMPL
O1054-03 - 1	2016	805	0.1840	-	7470A	PPB	10 Jan 2023 15:22:53	U	SMPL
O1054-04 - 1	2029	-124	-0.0136	-	7470A	PPB	10 Jan 2023 15:25:09	U	SMPL
CCV35 - 1	CCV35	25204	5.3729	-	7470A	PPB	10 Jan 2023 15:27:24	U	SMPL
CCB35 - 1	CCB35	-5	0.0117	-	7470A	PPB	10 Jan 2023 15:31:54	U	SMPL
O1065-02 - 1	VNJ-210	32	0.0196	-	7470A	PPB	10 Jan 2023 15:34:09	U	SMPL
O1069-02 - 1	TP-7	27	0.0185	-	7470A	PPB	10 Jan 2023 15:36:24	U	SMPL
O1069-02DUP - 1	TP-7DUP	30	0.0192	-	7470A	PPB	10 Jan 2023 15:38:40	U	SMPL
O1069-02MS - 1	TP-7MS	15527	3.3149	-	7470A	PPB	10 Jan 2023 15:40:55	U	SMPL
O1069-02MSD - 1	TP-7MSD	16302	3.4797	-	7470A	PPB	10 Jan 2023 15:43:11	U	SMPL
PB150153BL - 1	PBW	2	0.0132	-	7470A	PPB	10 Jan 2023 15:45:28	U	SMPL
O1054-05 - 1	TOTE-COMP	30	0.0192	-	7470A	PPB	10 Jan 2023 15:50:02	U	SMPL
O1054-06 - 1	COMP-03	210	0.0575	-	7470A	PPB	10 Jan 2023 15:52:19	U	SMPL
O1054-07 - 1	2026	17	0.0164	-	7470A	PPB	10 Jan 2023 15:54:35	U	SMPL
CCV36 - 1	CCV36	25561	5.4488	-	7470A	PPB	10 Jan 2023 15:56:50	U	SMPL
CCB36 - 1	CCB36	-35	0.0053	-	7470A	PPB	10 Jan 2023 15:59:06	U	SMPL
O1054-07DUP - 1	2026DUP	23	0.0177	-	7470A	PPB	10 Jan 2023 16:01:22	U	SMPL
O1054-07MS - 1	2026MS	16464	3.5141	-	7470A	PPB	10 Jan 2023 16:03:38	U	SMPL
O1054-07MSD - 1	2026MSD	15313	3.2694	-	7470A	PPB	10 Jan 2023 16:05:53	U	SMPL
PB150164BS - 1	LCSW	18839	4.0192	-	7470A	PPB	10 Jan 2023 16:10:24	U	SMPL
O1077-01 - 1	BUR-1550	1103	0.2474	-	7470A	PPB	10 Jan 2023 16:12:41	U	SMPL
O1081-04 - 1	599	40327	8.5890	-	7470A	PPB	10 Jan 2023 16:14:57	U	SMPL
O1081-05 - 1	601	515	0.1223	-	7470A	PPB	10 Jan 2023 16:17:14	U	SMPL
O1081-06 - 1	31944	4210	0.9081	-	7470A	PPB	10 Jan 2023 16:19:32	U	SMPL
O1081-07 - 1	31945	1452	0.3216	-	7470A	PPB	10 Jan 2023 16:21:47	U	SMPL
CCV37 - 1	CCV37	23501	5.0107	-	7470A	PPB	10 Jan 2023 16:24:03	U	SMPL
CCB37 - 1	CCB37	-73	-0.0027	-	7470A	PPB	10 Jan 2023 16:26:20	U	SMPL
O1081-08 - 1	31956	464	0.1115	-	7470A	PPB	10 Jan 2023 16:28:36	U	SMPL
O1082-02DLX5 - 1	1556	9598	2.0540	-	7470A	PPB	10 Jan 2023 16:30:51	U	SMPL
O1082-03 - 1	585	5118	1.1012	-	7470A	PPB	10 Jan 2023 16:33:06	U	SMPL
O1082-05 - 1	1518	213	0.0581	-	7470A	PPB	10 Jan 2023 16:35:22	U	SMPL
O1086-05 - 1	AUD-1843	411	0.1002	-	7470A	PPB	10 Jan 2023 16:37:38	U	SMPL
O1086-06 - 1	AUD-1844	552	0.1302	-	7470A	PPB	10 Jan 2023 16:39:54	U	SMPL
O1086-07 - 1	AUD-1845	291	0.0747	-	7470A	PPB	10 Jan 2023 16:42:11	U	SMPL
O1086-07DUP - 1	AUD-1845DUP	149	0.0445	-	7470A	PPB	10 Jan 2023 16:44:28	U	SMPL
O1086-07MS - 1	AUD-1845MS	15470	3.3028	-	7470A	PPB	10 Jan 2023 16:46:45	U	SMPL
O1086-07MSD - 1	AUD-1845MSD	16799	3.5854	-	7470A	PPB	10 Jan 2023 16:49:01	U	SMPL
CCV38 - 1	CCV38	23796	5.0734	-	7470A	PPB	10 Jan 2023 16:51:17	U	SMPL
CCB38 - 1	CCB38	-93	-0.0070	-	7470A	PPB	10 Jan 2023 16:53:34	U	SMPL
PB150105TB - 1	PB150105TB	36	0.0204	-	7470A	PPB	10 Jan 2023 16:55:49	U	SMPL
PB150107TB - 1	PB150107TB	48	0.0230	-	7470A	PPB	10 Jan 2023 16:58:05	U	SMPL
O1069-02LX5 - 1		-10	0.0107	-	7470A	PPB	10 Jan 2023 17:00:21	U	SMPL
O1069-02A - 1		17762	3.7902	-	7470A	PPB	10 Jan 2023 17:02:36	U	SMPL
O1054-07LX5 - 1		-49	0.0024	-	7470A	PPB	10 Jan 2023 17:04:52	U	SMPL
O1054-07A - 1		17650	3.7664	-	7470A	PPB	10 Jan 2023 17:07:09	U	SMPL
O1086-07LX5 - 1		55	0.0245	-	7470A	PPB	10 Jan 2023 17:09:25	U	SMPL
O1086-07A - 1		17719	3.7810	-	7470A	PPB	10 Jan 2023 17:11:42	U	SMPL
PB150153BS - 1	LCSW	18710	3.9918	-	7470A	PPB	10 Jan 2023 17:13:59	U	SMPL
PB150164BL - 1	PBW	-30	0.0064	-	7470A	PPB	10 Jan 2023 17:16:16	U	SMPL
CCV39 - 1	CCV39	23824	5.0794	-	7470A	PPB	10 Jan 2023 17:18:32	U	SMPL
CCB39 - 1	CCB39	-29	0.0066	-	7470A	PPB	10 Jan 2023 17:20:48	U	SMPL