



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB123616

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	01/10/2023	13:15
2	CAL2	1	Water	NA	NA	20.8	7.00	01/10/2023	13:16
3	CAL3	1	Water	NA	NA	20.8	10.03	01/10/2023	13:18
4	ICV	1	Water	NA	NA	20.7	6.98	01/10/2023	13:20
5	CCV1	1	Water	NA	NA	20.8	2.01	01/10/2023	15:05
6	O1087-01	1	Water	NA	NA	20.2	5.44	01/10/2023	15:10
7	O1095-01	1	Water	NA	NA	20.4	4.55	01/10/2023	15:20
8	O1095-04	1	Water	NA	NA	20.7	7.55	01/10/2023	15:25
9	O1095-04DUP	1	Water	NA	NA	20.8	7.56	01/10/2023	15:26
10	CCV2	1	Water	NA	NA	20.8	12.03	01/10/2023	15:35

WORKLIST(Hardcopy Internal Chain)

VB 123616

WorkList Name : ph o1095 WorkList ID : 166476 Department : Wet-Chemistry Date : 01-10-2023 14:58:53

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/04/2023	Water	O1087-01	pH	Cool 4 deg C	STER11		POOL-WATER	01/04/2023	9040C
01/09/2023	Water	O1095-01	pH	Cool 4 deg C	STER07	N11	CELL-A-POOL-WATER	01/09/2023	9040C
01/09/2023	Water	O1095-04	pH	Cool 4 deg C	STER07	N11	CONTROL	01/09/2023	9040C

Date/Time 01.10.23 15:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 01.10.23 16:10
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]