



Analytical Summary Report

Analysis Method: 1010B
Parameter: Flash Point
Run Number: LB123626
Thermometer ID: FLASHPOINT

Reviewed By: RUBINA
Supervisor Review By: Iwona
Ambient Barometric Pressure (mmHg): 771.00
Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	01/11/2023	10:00
2	O1081-07		1	12	100.00	>212.0	>212.0	01/11/2023	10:30
3	O1081-07DUP		1	12	100.00	>212.0	>212.0	01/11/2023	11:00
4	O1081-08		1	14	100.00	>212.0	>212.0	01/11/2023	11:30
5	O1082-02		1	13	100.00	>212.0	>212.0	01/11/2023	12:00
6	O1082-03		1	14	100.00	>212.0	>212.0	01/11/2023	12:30
7	O1082-04		1	13	100.00	>212.0	>212.0	01/11/2023	13:00
8	O1086-05		1	14	100.00	>212.0	>212.0	01/11/2023	13:30
9	O1086-06		1	12	100.00	>212.0	>212.0	01/11/2023	14:00
10	O1086-07		1	12	100.00	>212.0	>212.0	01/11/2023	14:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp1-11

WorkList ID : 166492

Department : Wet-Chemistry

Date : 01-11-2023 08:45:42

LB123626

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/12/2023	Water	O1081-07	Flash Point	Cool 4 deg C	PSEG03	N32	31945	01/09/2023	1010B
01/12/2023	Water	O1081-08	Flash Point	Cool 4 deg C	PSEG03	N32	31956	01/09/2023	1010B
01/12/2023	Water	O1082-02	Flash Point	Cool 4 deg C	PSEG03	N33	1556	01/09/2023	1010B
01/12/2023	Water	O1082-03	Flash Point	Cool 4 deg C	PSEG03	N33	585	01/09/2023	1010B
01/12/2023	Water	O1082-04	Flash Point	Cool 4 deg C	PSEG03	N33	31771	01/09/2023	1010B
01/12/2023	Water	O1086-05	Flash Point	Cool 4 deg C	PSEG03	N33	AUD-1843	01/09/2023	1010B
01/12/2023	Water	O1086-06	Flash Point	Cool 4 deg C	PSEG03	N33	AUD-1844	01/09/2023	1010B
01/12/2023	Water	O1086-07	Flash Point	Cool 4 deg C	PSEG03	N33	AUD-1845	01/09/2023	1010B

Date/Time 01/11/2023 09:30
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RS CEA-(26)

Date/Time 01/11/2023 15:30
 Raw Sample Received by: RS CEA-(26)
 Raw Sample Relinquished by: RM CWC