

LIB123627

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : 12

Instrument ID : Konelab

1/11/2023 10:52

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	1.006	0.0	0.108	
ICB1	-0.003	0.0	0.011	
CCV1	1.003	0.0	0.108	
CCB1	-0.002	0.0	0.011	
PB150168BL	0.000	0.0	0.011	
PB150168BS	1.001	0.0	0.107	
O1079-01	-0.000	0.0	0.011	
O1079-01DUP	-0.005	0.0	0.010	
O1079-01MS	1.138	0.0	0.121	
O1079-01MSD	1.139	0.0	0.121	
O1079-02	-0.000	0.0	0.011	
CCV2	1.003	0.0	0.108	
CCB2	-0.002	0.0	0.011	

N 13
Mean 0.483
SD 0.5467
CV% 113.19

Aquakem v. 7.2AQ1

Results from time period:

Wed Jan 11 09:46:04 2023

Wed Jan 11 10:49:19 2023

Sample Id	Sam/Ctr/c	Test short	Test type	Result	Result unit	Result date and time	Stat
0.00PPM	A	Phenolics-	P	-0.0022	mg/l	1/11/2023 9:46:04	
0.05PPM	A	Phenolics-	P	0.0537	mg/l	1/11/2023 9:46:05	
0.1PPM	A	Phenolics-	P	0.1073	mg/l	1/11/2023 9:46:06	
0.25PPM	A	Phenolics-	P	0.2542	mg/l	1/11/2023 9:46:07	
0.50PPM	A	Phenolics-	P	0.4996	mg/l	1/11/2023 9:46:08	
1.0PPM	A	Phenolics-	P	0.9765	mg/l	1/11/2023 9:46:09	
2.0PPM	A	Phenolics-	P	2.0109	mg/l	1/11/2023 9:46:10	
ICV1	S	Phenolics-	P	1.0056	mg/l	1/11/2023 10:42:27	
ICB1	S	Phenolics-	P	-0.0025	mg/l	1/11/2023 10:42:28	
CCV1	S	Phenolics-	P	1.0033	mg/l	1/11/2023 10:42:31	
CCB1	S	Phenolics-	P	-0.0022	mg/l	1/11/2023 10:42:33	
PB150168BL	S	Phenolics-	P	0.0001	mg/l	1/11/2023 10:42:35	
PB150168BS	S	Phenolics-	P	1.0013	mg/l	1/11/2023 10:42:37	
O1079-01	S	Phenolics-	P	-0.0001	mg/l	1/11/2023 10:49:12	
O1079-01DUP	S	Phenolics-	P	-0.0046	mg/l	1/11/2023 10:49:13	
O1079-01MS	S	Phenolics-	P	1.138	mg/l	1/11/2023 10:49:14	
O1079-01MSD	S	Phenolics-	P	1.139	mg/l	1/11/2023 10:49:15	
O1079-02	S	Phenolics-	P	-0.0001	mg/l	1/11/2023 10:49:16	
CCV2	S	Phenolics-	P	1.0032	mg/l	1/11/2023 10:49:17	
CCB2	S	Phenolics-	P	-0.0021	mg/l	1/11/2023 10:49:18	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

1/11/2023 9:46

Reviewed by : 12

Instrument ID : Konelab


Test Phenolics-

Accepted 1/11/2023 9:46

Factor 10.37
Bias 0.011

Coeff. of det. 0.999758

Errors



	Calibrator	Response	Calc. con.	Conc.	Re Errors
1	0.00PPM	0.011	-0.0022	0.0000	-
2	Phenol-2	0.016	0.0537	0.0500	7.4
3	Phenol-2	0.021	0.1073	0.1000	7.3
4	Phenol-2	0.035	0.2542	0.2500	1.7
5	Phenol-2	0.059	0.4996	0.5000	-2.1
6	Phenol-2	0.105	0.9765	1.0000	-2.4
7	Phenol-2	0.205	2.0109	2.0000	0.5

12/11/23