

LB123629


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

54973



A= 0.0000e+000

B= 1.8342e-004

C= -9.4932e-002

Rho= 0.9999897

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.003	0.003	535	0.000	535					
0.2	0.200	0.183	-0.017	1517	0.0 %	1517					
2.5	2.500	2.523	0.023	14273	0.0 %	14273					
5.0	5.000	4.982	-0.018	27679	0.0 %	27679					
7.5	7.500	7.520	0.020	41516	0.0 %	41516					
10.0	10.000	9.988	-0.012	54973	0.0 %	54973					

0.00
-9
1
0
0
0

LB123629 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 11 Jan 2023 11:46:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	535	-	0.0000	7471B	PPB	11 Jan 2023 11:57:14	S	Std
0.2 - 1	50.2	1517	-	0.2000	7471B	PPB	11 Jan 2023 11:59:36	S	Std
2.5 - 1	52.5	14273	-	2.5000	7471B	PPB	11 Jan 2023 12:01:53	S	Std
5.0 - 1	55	27679	-	5.0000	7471B	PPB	11 Jan 2023 12:04:10	S	Std
7.5 - 1	57.5	41516	-	7.5000	7471B	PPB	11 Jan 2023 12:06:26	S	Std
10.0 - 1	51.0	54973	-	10.0000	7471B	PPB	11 Jan 2023 12:08:43	S	Std
ICV99 - 1	ICV99	21265	3.8055	-	7471B	PPB	11 Jan 2023 12:11:27	U	SMPL
ICB99 - 1	ICB99	760	0.0445	-	7471B	PPB	11 Jan 2023 12:13:42	U	SMPL
CCV40 - 1	CCV40	27671	4.9806	-	7471B	PPB	11 Jan 2023 12:18:38	U	SMPL
CCB40 - 1	CCB40	743	0.0414	-	7471B	PPB	11 Jan 2023 12:20:53	U	SMPL
CRA - 1	CRA	1592	0.1971	-	7471B	PPB	11 Jan 2023 12:23:08	U	SMPL
HighStd - 1	HighStd	54469	9.8959	-	7471B	PPB	11 Jan 2023 12:25:23	U	SMPL
ChkStd - 1	ChkStd	38816	7.0248	-	7471B	PPB	11 Jan 2023 12:27:39	U	SMPL
PB150182BL - 1	PBS	609	0.0168	-	7471B	PPB	11 Jan 2023 12:29:54	U	SMPL
PB150182BS - 1	LCSS	19833	3.5429	-	7471B	PPB	11 Jan 2023 12:32:10	U	SMPL
O1072-01 - 1	H-5-MIDDLE(4-8)	1068	0.1010	-	7471B	PPB	11 Jan 2023 12:34:28	U	SMPL
O1076-01 - 1	BU-03-010923	1943	0.2615	-	7471B	PPB	11 Jan 2023 12:36:46	U	SMPL
O1083-01 - 1	OK-01-010923	8321	1.4313	-	7471B	PPB	11 Jan 2023 12:39:03	U	SMPL
O1084-01 - 1	MH-2	1030	0.0940	-	7471B	PPB	11 Jan 2023 12:41:21	U	SMPL
O1089-01 - 1	PL-1-010923	6625	1.1202	-	7471B	PPB	11 Jan 2023 12:43:36	U	SMPL
CCV41 - 1	CCV41	27722	4.9899	-	7471B	PPB	11 Jan 2023 12:45:52	U	SMPL
CCB41 - 1	CCB41	796	0.0511	-	7471B	PPB	11 Jan 2023 12:48:08	U	SMPL
O1089-03 - 1	PL-2-010923	9217	1.5957	-	7471B	PPB	11 Jan 2023 12:50:24	U	SMPL
O1089-05 - 1	PL-3-010923	6913	1.1731	-	7471B	PPB	11 Jan 2023 12:52:39	U	SMPL
O1089-05DUP - 1	PL-3-010923DUP	7824	1.3402	-	7471B	PPB	11 Jan 2023 12:59:50	U	SMPL
O1089-05MSD - 1	PL-3-010923MSD	29178	5.2570	-	7471B	PPB	11 Jan 2023 13:04:22	U	SMPL
O1090-01 - 1	SU-3-010923	1643	0.2064	-	7471B	PPB	11 Jan 2023 13:06:38	U	SMPL
O1093-01 - 1	TT-214-IDWSO-20230109-1	679	0.0296	-	7471B	PPB	11 Jan 2023 13:08:55	U	SMPL
O1094-01 - 1	STOCKPILE-A	1764	0.2286	-	7471B	PPB	11 Jan 2023 13:11:12	U	SMPL
O1094-03 - 1	STOCKPILE-B	1586	0.1960	-	7471B	PPB	11 Jan 2023 13:13:28	U	SMPL
O1106-01 - 1	MH-3	1106	0.1079	-	7471B	PPB	11 Jan 2023 13:15:44	U	SMPL
CCV42 - 1	CCV42	27839	5.0114	-	7471B	PPB	11 Jan 2023 13:18:00	U	SMPL
CCB42 - 1	CCB42	616	0.0181	-	7471B	PPB	11 Jan 2023 13:20:16	U	SMPL
O1106-04 - 1	TP-B	1762	0.2283	-	7471B	PPB	11 Jan 2023 13:22:31	U	SMPL
O1109-01 - 1	OR-03-011023	4508	0.7319	-	7471B	PPB	11 Jan 2023 13:24:46	U	SMPL
O1089-05MS - 1	PL-3-010923MS	29265	5.2729	-	7471B	PPB	11 Jan 2023 13:27:02	U	SMPL
O1089-05LX5 - 1		1122	0.1109	-	7471B	PPB	11 Jan 2023 13:29:18	U	SMPL
O1089-05A - 1		27398	4.9305	-	7471B	PPB	11 Jan 2023 13:31:34	U	SMPL
CCV43 - 1	CCV43	27436	4.9375	-	7471B	PPB	11 Jan 2023 13:33:50	U	SMPL
CCB43 - 1	CCB43	464	-0.0098	-	7471B	PPB	11 Jan 2023 13:36:06	U	SMPL