



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB123636

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	01/11/2023	14:15
2	CAL2	1	Water	NA	NA	20.8	7.00	01/11/2023	14:16
3	CAL3	1	Water	NA	NA	20.8	10.03	01/11/2023	14:18
4	ICV	1	Water	NA	NA	20.7	6.98	01/11/2023	14:20
5	CCV1	1	Water	NA	NA	20.7	2.01	01/11/2023	14:22
6	O1092-01	1	Water	NA	NA	22.3	7.47	01/11/2023	14:25
7	O1092-03	1	Water	NA	NA	2.3	7.19	01/11/2023	14:30
8	O1114-01	1	Water	NA	NA	20.1	8.55	01/11/2023	14:40
9	O1114-02	1	Water	NA	NA	20.1	8.40	01/11/2023	14:43
10	O1114-03	1	Water	NA	NA	20.2	8.90	01/11/2023	14:45
11	O1114-04	1	Water	NA	NA	20.1	8.26	01/11/2023	14:47
12	O1114-04DUP	1	Water	NA	NA	20.2	8.27	01/11/2023	14:48
13	CCV2	1	Water	NA	NA	20.7	12.02	01/11/2023	14:50

W312363630
W3123636

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O1114

WorkList ID : 166510

Department : Wet-Chemistry

Date : 01-11-2023 14:03:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/10/2023	Water	O1092-01	pH	A	Cool 4 deg C	PSEG03	N41 FMI-120	01/10/2023	9040C
01/10/2023	Water	O1092-03	pH	A	Cool 4 deg C	PSEG03	N41 RL-TOTE-0110	01/10/2023	9040C
01/11/2023	Water	O1114-01	pH	A	Cool 4 deg C	PSEG03	N41 HARRISON-MR-1	01/11/2023	9040C
01/11/2023	Water	O1114-02	pH	A	Cool 4 deg C	PSEG03	N41 HARRISON-MR-2	01/11/2023	9040C
01/11/2023	Water	O1114-03	pH	A	Cool 4 deg C	PSEG03	N41 HARRISON-MR-3	01/11/2023	9040C
01/11/2023	Water	O1114-04	pH	A	Cool 4 deg C	PSEG03	N41 HARRISON-MR-4	01/11/2023	9040C

Date/Time 01-11-23 16:10

Raw Sample Received by: ZP (W3C)

Raw Sample Relinquished by: ZP (W3C)

Date/Time 01-11-23 15:20

Raw Sample Received by: ZP (W3C)

Raw Sample Relinquished by: ZP (W3C)