

LB123638


245.1

INSTRUMENT ID: CV4

Linear

 $\mu$  Abs.:

42552



A= 0.0000e+000

B= 2.3690e-004

C= -9.7516e-003

Rho= 0.9998663

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | U/D |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|-----|
| 0.0    | 0.000  | 0.002  | 0.002  | 51    | 0.000      | 51    |       |       |       |       | 12  |
| 0.05   | 0.050  | 0.056  | 0.006  | 276   | 0.0 %      | 276   | 0     |       |       |       | 10  |
| 0.2    | 0.200  | 0.180  | -0.020 | 802   | 0.0 %      | 802   |       |       |       |       | 2   |
| 2.5    | 2.500  | 2.542  | 0.042  | 10770 | 0.0 %      | 10770 |       |       |       |       | 1   |
| 5.0    | 5.000  | 5.032  | 0.032  | 21281 | 0.0 %      | 21281 |       |       |       |       | 2   |
| 7.5    | 7.500  | 7.367  | -0.133 | 31140 | 0.0 %      | 31140 |       |       |       |       | 1   |
| 10.0   | 10.000 | 10.071 | 0.071  | 42552 | 0.0 %      | 42552 |       |       |       |       | 1   |

## LB123638 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 11 Jan 2023 14:46:35

| Sample ID       | Extended ID       | μ Abs. | Conc.   | Std Conc | Method | Units | Date                 | Type | Type |
|-----------------|-------------------|--------|---------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | 50                | 51     | -       | 0.0000   | 245.1  | PPB   | 11 Jan 2023 14:49:36 | S    | Std  |
| 0.05 - 1        | 50.250.05         | 276    | -       | 0.0500   | 245.1  | PPB   | 11 Jan 2023 14:51:54 | S    | Std  |
| 0.2 - 1         | 52.5              | 802    | -       | 0.2000   | 245.1  | PPB   | 11 Jan 2023 14:54:11 | S    | Std  |
| 2.5 - 1         | 55                | 10770  | -       | 2.5000   | 245.1  | PPB   | 11 Jan 2023 14:56:28 | S    | Std  |
| 5.0 - 1         | 57.5              | 21281  | -       | 5.0000   | 245.1  | PPB   | 11 Jan 2023 14:58:45 | S    | Std  |
| 7.5 - 1         | 510               | 31140  | -       | 7.5000   | 245.1  | PPB   | 11 Jan 2023 15:01:00 | S    | Std  |
| 10.0 - 1        |                   | 42552  | -       | 10.0000  | 245.1  | PPB   | 11 Jan 2023 15:06:54 | S    | Std  |
| ICV100 - 1      | ICV100            | 16928  | 4.0005  | -245.1   |        | PPB   | 11 Jan 2023 15:12:11 | U    | SMPL |
| ICB100 - 1      | ICB100            | 35     | -0.0015 | -245.1   |        | PPB   | 11 Jan 2023 15:14:26 | U    | SMPL |
| CCV44 - 1       | CCV44             | 21311  | 5.0389  | -245.1   |        | PPB   | 11 Jan 2023 15:16:41 | U    | SMPL |
| CCB44 - 1       | CCB44             | 2      | -0.0093 | -245.1   |        | PPB   | 11 Jan 2023 15:18:56 | U    | SMPL |
| CRA - 1         | CRA               | 297    | 0.0606  | -245.1   |        | PPB   | 11 Jan 2023 15:21:11 | U    | SMPL |
| HighStd - 1     | HighStd           | 42844  | 10.1401 | -245.1   |        | PPB   | 11 Jan 2023 15:23:26 | U    | SMPL |
| ChkStd - 1      | ChkStd            | 29793  | 7.0483  | -245.1   |        | PPB   | 11 Jan 2023 15:25:42 | U    | SMPL |
| PB150191BL - 1  | PBW               | 13     | -0.0067 | -245.1   |        | PPB   | 11 Jan 2023 15:27:57 | U    | SMPL |
| PB150191BS - 1  | LCSW              | 15771  | 3.7264  | -245.1   |        | PPB   | 11 Jan 2023 15:30:13 | U    | SMPL |
| O1055-01 - 1    | MH-J              | 198    | 0.0372  | -245.1   |        | PPB   | 11 Jan 2023 15:32:30 | U    | SMPL |
| O1055-01DUP - 1 | MH-JDUP           | 218    | 0.0419  | -245.1   |        | PPB   | 11 Jan 2023 15:34:46 | U    | SMPL |
| O1055-01MS - 1  | MH-JMS            | 17160  | 4.0555  | -245.1   |        | PPB   | 11 Jan 2023 15:37:03 | U    | SMPL |
| O1055-01MSD - 1 | MH-JMSD           | 14755  | 3.4857  | -245.1   |        | PPB   | 11 Jan 2023 15:39:21 | U    | SMPL |
| O1055-02 - 1    | MH-J              | 45     | 0.0009  | -245.1   |        | PPB   | 11 Jan 2023 15:41:36 | U    | SMPL |
| CCV45 - 1       | CCV45             | 21449  | 5.0716  | -245.1   |        | PPB   | 11 Jan 2023 15:43:51 | U    | SMPL |
| CCB45 - 1       | CCB45             | 8      | -0.0079 | -245.1   |        | PPB   | 11 Jan 2023 15:46:07 | U    | SMPL |
| O1055-03 - 1    | MH-F              | 216    | 0.0414  | -245.1   |        | PPB   | 11 Jan 2023 15:48:22 | U    | SMPL |
| O1055-04 - 1    | MH-F              | 110    | 0.0163  | -245.1   |        | PPB   | 11 Jan 2023 15:50:37 | U    | SMPL |
| O1055-05 - 1    | MH-G              | 104    | 0.0149  | -245.1   |        | PPB   | 11 Jan 2023 15:52:53 | U    | SMPL |
| O1055-06 - 1    | MH-G              | 94     | 0.0125  | -245.1   |        | PPB   | 11 Jan 2023 15:55:09 | U    | SMPL |
| O1079-01 - 1    | SAMPLE-1-EFFLUENT | 95     | 0.0128  | -245.1   |        | PPB   | 11 Jan 2023 15:57:25 | U    | SMPL |
| O1079-02 - 1    | SAMPLE-2-INFLUENT | 222    | 0.0428  | -245.1   |        | PPB   | 11 Jan 2023 15:59:41 | U    | SMPL |
| O1055-01LX5 - 1 |                   | 168    | 0.0300  | -245.1   |        | PPB   | 11 Jan 2023 16:01:58 | U    | SMPL |
| O1055-01A - 1   |                   | 17107  | 4.0429  | -245.1   |        | PPB   | 11 Jan 2023 16:04:15 | U    | SMPL |
| CCV46 - 1       | CCV46             | 21650  | 5.1192  | -245.1   |        | PPB   | 11 Jan 2023 16:06:33 | U    | SMPL |
| CCB46 - 1       | CCB46             | 9      | -0.0076 | -245.1   |        | PPB   | 11 Jan 2023 16:08:48 | U    | SMPL |