



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123649

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	01/12/2023	10:00
2	O1092-01		1	12	100.00	>212.0	>212.0	01/12/2023	10:30
3	O1092-01DUP		1	12	100.00	>212.0	>212.0	01/12/2023	11:00
4	O1092-03		1	14	100.00	>212.0	>212.0	01/12/2023	11:30
5	O1103-01		1	12	100.00	>212.0	>212.0	01/12/2023	12:00
6	O1111-04		1	13	100.00	>212.0	>212.0	01/12/2023	12:30
7	O1111-05		1	14	100.00	>212.0	>212.0	01/12/2023	13:00
8	O1111-06		1	12	100.00	>212.0	>212.0	01/12/2023	13:30
9	O1111-07		1	14	100.00	>212.0	>212.0	01/12/2023	14:00
10	O1111-08		1	14	100.00	>212.0	>212.0	01/12/2023	14:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66123649

WorkList Name : fp1-12

WorkList ID : 166512

Department : Wet-Chemistry

Date : 01-12-2023 08:13:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/17/2023	Water	O1092-01	Flash Point	Cool 4 deg C	PSEG03	N41	FMI-120	01/10/2023	1010B
01/17/2023	Water	O1092-03	Flash Point	Cool 4 deg C	PSEG03	N41	RL-TOTE-0110	01/10/2023	1010B
01/17/2023	Water	O1103-01	Flash Point	Cool 4 deg C	PSEG03	N41	332	01/10/2023	1010B
01/17/2023	Water	O1111-04	Flash Point	Cool 4 deg C	PSEG03	N41	D2806	01/10/2023	1010B
01/17/2023	Water	O1111-05	Flash Point	Cool 4 deg C	PSEG03	N41	D2838	01/10/2023	1010B
01/17/2023	Water	O1111-06	Flash Point	Cool 4 deg C	PSEG03	N41	T2867	01/10/2023	1010B
01/17/2023	Water	O1111-07	Flash Point	Cool 4 deg C	PSEG03	N41	T2868	01/10/2023	1010B
01/17/2023	Water	O1111-08	Flash Point	Cool 4 deg C	PSEG03	N41	T2733	01/10/2023	1010B

Date/Time 01/12/2023 09:50
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: SP (CWC)

Date/Time 01/12/2023 14:45
 Raw Sample Received by: SP (CWC)
 Raw Sample Relinquished by: RM (CWC)