

L8123657


7470A

INSTRUMENT ID: CV2

Linear

μ Abs.:

4709E



A= 0.0000e+000

B= 2.1582e-004

C= -6.8603e-002

Rho= 0.9998137

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.041	0.041	506	0.000	506					
0.2	0.200	0.185	-0.015	1177	0.0 %	1177					
2.5	2.500	2.512	0.012	11956	0.0 %	11956					
5.0	5.000	5.005	0.005	23509	0.0 %	23509					
7.5	7.500	7.361	-0.139	34426	0.0 %	34426					
10.0	10.000	10.096	0.096	47096	0.0 %	47096					

0.001
1.002
1.8

LB123657 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 12 Jan 2023 12:09:54

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	506	-	0.0000	7470A	PPB	12 Jan 2023 12:17:32	S	Std
0.2 - 1	51.2	1177	-	0.2000	7470A	PPB	12 Jan 2023 12:19:49	S	Std
2.5 - 1	52.5	11956	-	2.5000	7470A	PPB	12 Jan 2023 12:22:06	S	Std
5.0 - 1	53	23509	-	5.0000	7470A	PPB	12 Jan 2023 12:24:23	S	Std
7.5 - 1	57.8	34426	-	7.5000	7470A	PPB	12 Jan 2023 12:26:39	S	Std
10.0 - 1	59.0	47096	-	10.0000	7470A	PPB	12 Jan 2023 12:31:29	S	Std
ICV101 - 1	ICV101	18068	3.8309	-	7470A	PPB	12 Jan 2023 12:34:12	U	SMPL
ICB101 - 1	ICB101	311	-0.0015	-	7470A	PPB	12 Jan 2023 12:36:26	U	SMPL
CCV47 - 1	CCV47	23187	4.9357	-	7470A	PPB	12 Jan 2023 12:38:42	U	SMPL
CCB47 - 1	CCB47	266	-0.0112	-	7470A	PPB	12 Jan 2023 12:40:57	U	SMPL
CRA - 1	CRA	1217	0.1941	-	7470A	PPB	12 Jan 2023 12:43:12	U	SMPL
HighStd - 1	HighStd	47382	10.1575	-	7470A	PPB	12 Jan 2023 12:45:27	U	SMPL
ChkStd - 1	ChkStd	32443	6.9333	-	7470A	PPB	12 Jan 2023 12:47:43	U	SMPL
PB150202BL - 1	PBW	361	0.0093	-	7470A	PPB	12 Jan 2023 12:49:59	U	SMPL
PB150202BS - 1	LCSW	18343	3.8902	-	7470A	PPB	12 Jan 2023 12:52:15	U	SMPL
O1084-02 - 1	MH-2	390	0.0156	-	7470A	PPB	12 Jan 2023 12:54:31	U	SMPL
O1084-02DUP - 1	MH-2DUP	406	0.0190	-	7470A	PPB	12 Jan 2023 12:56:47	U	SMPL
O1084-02MS - 1	MH-2MS	17022	3.6051	-	7470A	PPB	12 Jan 2023 12:59:04	U	SMPL
O1084-02MSD - 1	MH-2MSD	17475	3.7029	-	7470A	PPB	12 Jan 2023 13:01:21	U	SMPL
O1106-02 - 1	MH-3	397	0.0171	-	7470A	PPB	12 Jan 2023 13:03:37	U	SMPL
CCV48 - 1	CCV48	24224	5.1595	-	7470A	PPB	12 Jan 2023 13:05:52	U	SMPL
CCB48 - 1	CCB48	328	0.0022	-	7470A	PPB	12 Jan 2023 13:08:07	U	SMPL
O1106-05 - 1	TP-B	575	0.0555	-	7470A	PPB	12 Jan 2023 13:10:22	U	SMPL
O1108-01 - 1	2739	429	0.0240	-	7470A	PPB	12 Jan 2023 13:12:37	U	SMPL
O1111-01 - 1	D2856	412	0.0203	-	7470A	PPB	12 Jan 2023 13:14:53	U	SMPL
O1111-02 - 1	D2872	363	0.0097	-	7470A	PPB	12 Jan 2023 13:17:08	U	SMPL
O1111-03 - 1	D2873	422	0.0225	-	7470A	PPB	12 Jan 2023 13:19:24	U	SMPL
PB150203BL - 1	PBW	413	0.0205	-	7470A	PPB	12 Jan 2023 13:21:41	U	SMPL
PB150203BS - 1	LCSW	16775	3.5518	-	7470A	PPB	12 Jan 2023 13:23:57	U	SMPL
O1103-01 - 1	332	367	0.0106	-	7470A	PPB	12 Jan 2023 13:26:14	U	SMPL
O1103-01DUP - 1	332DUP	428	0.0238	-	7470A	PPB	12 Jan 2023 13:28:31	U	SMPL
O1103-01MS - 1	332MS	17131	3.6286	-	7470A	PPB	12 Jan 2023 13:30:46	U	SMPL
CCV49 - 1	CCV49	24457	5.2098	-	7470A	PPB	12 Jan 2023 13:33:02	U	SMPL
CCB49 - 1	CCB49	266	-0.0112	-	7470A	PPB	12 Jan 2023 13:35:17	U	SMPL
O1103-01MSD - 1	332MSD	17207	3.6450	-	7470A	PPB	12 Jan 2023 13:40:11	U	SMPL
O1111-04 - 1	D2806	404	0.0186	-	7470A	PPB	12 Jan 2023 13:42:26	U	SMPL
O1111-05 - 1	D2838	413	0.0205	-	7470A	PPB	12 Jan 2023 13:44:42	U	SMPL
O1111-06 - 1	T2867	416	0.0212	-	7470A	PPB	12 Jan 2023 13:46:57	U	SMPL
O1111-07 - 1	T2868	430	0.0242	-	7470A	PPB	12 Jan 2023 13:49:13	U	SMPL
O1111-08 - 1	T2733	434	0.0251	-	7470A	PPB	12 Jan 2023 13:51:30	U	SMPL
PB150154TB - 1	PB150154TB	476	0.0341	-	7470A	PPB	12 Jan 2023 13:53:46	U	SMPL
PB150167TB - 1	PB150167TB	402	0.0182	-	7470A	PPB	12 Jan 2023 13:56:03	U	SMPL
CCV50 - 1	CCV50	22501	4.7876	-	7470A	PPB	12 Jan 2023 14:11:35	U	SMPL
CCB50 - 1	CCB50	228	-0.0194	-	7470A	PPB	12 Jan 2023 14:13:50	U	SMPL
O1084-02LX5 - 1		-16	-0.0721	-	7470A	PPB	12 Jan 2023 14:16:05	U	SMPL
O1084-02A - 1		16960	3.5917	-	7470A	PPB	12 Jan 2023 14:21:01	U	SMPL
O1103-01LX5 - 1		-18	-0.0725	-	7470A	PPB	12 Jan 2023 14:23:16	U	SMPL
O1103-01A - 1		16214	3.4307	-	7470A	PPB	12 Jan 2023 14:25:31	U	SMPL
CCV51 - 1	CCV51	22310	4.7464	-	7470A	PPB	12 Jan 2023 14:27:47	U	SMPL
CCB51 - 1	CCB51	240	-0.0168	-	7470A	PPB	12 Jan 2023 14:30:03	U	SMPL