

LB123662


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

53084



A= 0.0000e+000

B= 1.8755e-004

C= 6.8840e-004

Rho= 0.9998925

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.027	0.027	140	0.000	140					
0.2	0.200	0.204	0.004	1084	0.0 %	1084					2
2.5	2.500	2.405	-0.095	12820	0.0 %	12820					-4
5.0	5.000	5.061	0.061	26981	0.0 %	26981					1
7.5	7.500	7.547	0.047	40236	0.0 %	40236					1
10.0	10.000	9.956	-0.044	53084	0.0 %	53084					0

LB123662 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 Jan 2023 15:11:05

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	140	-	0.0000	7471B	PPB	12 Jan 2023 15:56:05	S	Std
0.2 - 1	50.2	1084	-	0.2000	7471B	PPB	12 Jan 2023 15:58:28	S	Std
2.5 - 1	52.5	12820	-	2.5000	7471B	PPB	12 Jan 2023 16:00:47	S	Std
5.0 - 1	55	26981	-	5.0000	7471B	PPB	12 Jan 2023 16:03:04	S	Std
7.5 - 1	57.5	40236	-	7.5000	7471B	PPB	12 Jan 2023 16:05:21	S	Std
10.0 - 1	59	53084	-	10.0000	7471B	PPB	12 Jan 2023 16:07:37	S	Std
ICV01 - 1	ICV01	20881	3.9168	-	7471B	PPB	12 Jan 2023 16:10:20	U	SMPL
ICB01 - 1	ICB01	149	0.0286	-	7471B	PPB	12 Jan 2023 16:12:35	U	SMPL
CCV52 - 1	CCV52	27131	5.0890	-	7471B	PPB	12 Jan 2023 16:14:50	U	SMPL
CCB52 - 1	CCB52	121	0.0234	-	7471B	PPB	12 Jan 2023 16:17:05	U	SMPL
CRA - 1	CRA	1118	0.2104	-	7471B	PPB	12 Jan 2023 16:19:19	U	SMPL
HighStd - 1	HighStd	53352	10.0066	-	7471B	PPB	12 Jan 2023 16:21:34	U	SMPL
ChkStd - 1	ChkStd	37280	6.9924	-	7471B	PPB	12 Jan 2023 16:23:50	U	SMPL
PB150228BL - 1	PBS	74	0.0146	-	7471B	PPB	12 Jan 2023 16:26:06	U	SMPL
PB150228BS - 1	LCSS	17726	3.3251	-	7471B	PPB	12 Jan 2023 16:28:24	U	SMPL
O1122-01 - 1	PAT-03-011123	6501	1.2199	-	7471B	PPB	12 Jan 2023 16:30:40	U	SMPL
O1124-01 - 1	TP-C	4942	0.9275	-	7471B	PPB	12 Jan 2023 16:32:56	U	SMPL
O1124-01DUP - 1	TP-CDUP	5397	1.0129	-	7471B	PPB	12 Jan 2023 16:38:06	U	SMPL
O1124-01MS - 1	TP-CMS	27376	5.1349	-	7471B	PPB	12 Jan 2023 16:40:23	U	SMPL
O1124-01MSD - 1	TP-CMSD	27136	5.0899	-	7471B	PPB	12 Jan 2023 16:47:27	U	SMPL
CCV53 - 1	CCV53	27215	5.1048	-	7471B	PPB	12 Jan 2023 16:49:42	U	SMPL
CCB53 - 1	CCB53	80	0.0157	-	7471B	PPB	12 Jan 2023 16:51:57	U	SMPL
O1124-04 - 1	MH-4	1210	0.2276	-	7471B	PPB	12 Jan 2023 16:54:12	U	SMPL
O1125-01 - 1	B-04-SB01	18208	3.4155	-	7471B	PPB	12 Jan 2023 16:56:29	U	SMPL
O1124-01LX5 - 1		1048	0.1972	-	7471B	PPB	12 Jan 2023 16:58:44	U	SMPL
O1124-01A - 1		26746	5.0168	-	7471B	PPB	12 Jan 2023 17:01:00	U	SMPL
CCV54 - 1	CCV54	27173	5.0969	-	7471B	PPB	12 Jan 2023 17:03:16	U	SMPL
CCB54 - 1	CCB54	64	0.0127	-	7471B	PPB	12 Jan 2023 17:05:32	U	SMPL