



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123676

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	01/16/2023	10:00
2	O1115-02		1	12	100.00	>212.0	>212.0	01/16/2023	10:30
3	O1115-02DUP		1	14	100.00	>212.0	>212.0	01/16/2023	11:00
4	O1155-02		1	13	100.00	>212.0	>212.0	01/16/2023	11:30
5	O1155-04		1	14	100.00	>212.0	>212.0	01/16/2023	12:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6123676

WorkList Name : fp-s-1-13

WorkList ID : 166559

Department : Wet-Chemistry

Date : 01-13-2023 10:05:10

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/18/2023	Solid	O1115-02	Flash Point	Cool 4 deg C	PSEG03	N42	SAS-OIL	01/11/2023	1010B
01/18/2023	Solid	O1155-02	Flash Point	Cool 4 deg C	PSEG03	N51	33896	01/13/2023	1010B
01/18/2023	Solid	O1155-04	Flash Point	Cool 4 deg C	PSEG03	N51	31532	01/13/2023	1010B

Date/Time 01/16/2023 09:30
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: JWC

Date/Time 01/16/2023 12:40
 Raw Sample Received by: JWC
 Raw Sample Relinquished by: RM CWC