



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123677

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.7	01/13/2023	11:10
2	O1114-01		1	12	100.00	>212.0	>212.0	01/13/2023	11:40
3	O1114-02		1	14	100.00	>212.0	>212.0	01/13/2023	12:10
4	O1114-03		1	12	100.00	>212.0	>212.0	01/13/2023	12:40
5	O1114-04		1	13	100.00	>212.0	>212.0	01/13/2023	13:10
6	O1134-02		1	12	100.00	>212.0	>212.0	01/13/2023	13:40
7	O1134-02DUP		1	14	100.00	>212.0	>212.0	01/13/2023	15:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6123677

WorkList Name : fp1-13

WorkList ID : 166548

Department : Wet-Chemistry

Date : 01-13-2023 08:08:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/18/2023	Water	O1114-01	Flash Point	Cool 4 deg C	PSEG03	N41	HARRISON-MR-1	01/11/2023	1010B
01/18/2023	Water	O1114-02	Flash Point	Cool 4 deg C	PSEG03	N41	HARRISON-MR-2	01/11/2023	1010B
01/18/2023	Water	O1114-03	Flash Point	Cool 4 deg C	PSEG03	N41	HARRISON-MR-3	01/11/2023	1010B
01/18/2023	Water	O1114-04	Flash Point	Cool 4 deg C	PSEG03	N41	HARRISON-MR-4	01/11/2023	1010B
01/17/2023	Water	O1134-02	Flash Point	Cool 4 deg C	PSEG03	N43	1115	01/12/2023	1010B

Date/Time 01/13/2023 11:00
 Raw Sample Received by: RM 'CWE'
 Raw Sample Relinquished by: RM 'CWE'

Date/Time 01/13/2023 15:50
 Raw Sample Received by: RM 'CWE'
 Raw Sample Relinquished by: RM 'CWE'