



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 1/16/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 14:40
In Date: 01/13/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 01/14/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB123682

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O1145-01	A43S6	1	1.12	8.49	9.61	8.91	91.8	
O1145-02	A43S7	2	1.16	8.54	9.7	9.11	93.1	
O1145-03	A43S7MS	3	1.16	8.54	9.7	9.11	93.1	
O1145-04	A43S7MSD	4	1.16	8.54	9.7	9.11	93.1	
O1145-05	A43T8	5	1.19	8.57	9.76	8.76	88.3	
O1145-06	A43T9	6	1.14	8.49	9.63	8.46	86.2	
O1145-07	A43W0	7	1.18	8.53	9.71	8.65	87.6	
O1145-08	A43W5	8	1.16	8.68	9.84	9.06	91.0	
O1145-09	A43W6	9	1.12	8.81	9.93	9.22	91.9	
O1145-10	A43X8	10	1.13	8.84	9.97	8.81	86.9	
O1145-11	A43X9	11	1.14	8.77	9.91	9.2	91.9	
O1145-12	A44A5	12	1.12	8.66	9.78	9.34	94.9	
O1145-13	A44A7	13	1.13	8.67	9.8	9.15	92.5	
O1145-14	A44A9	14	1.18	8.43	9.61	9.14	94.4	
O1145-15	A44B1	15	1.19	8.56	9.75	9.17	93.2	
O1145-16	A44B3	16	1.17	8.40	9.57	8.77	90.5	
O1145-17	A44B5	17	1.13	8.65	9.78	8.46	84.7	
O1145-18	A44B7	18	1.13	8.69	9.82	9.13	92.1	
O1145-19	A44B9	19	1.12	8.41	9.53	8.99	93.6	
O1145-20	A44C1	20	1.15	8.45	9.6	9.2	95.3	
O1145-21	A44C9	21	1.16	8.60	9.76	8.84	89.3	
O1145-22	A44D1	22	1.12	8.75	9.87	8.79	87.7	
O1145-23	A44E2	23	1.00	1.00	2.00	2.00	100.0	P.E. SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$