



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 1/16/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 16:15
In Date: 01/13/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:20
Out Date: 01/14/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB123685

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O1151-01	EXGP9	1	1.13	8.56	9.69	8.41	85.0	
O1151-02	EXGQ0	2	1.16	8.60	9.76	8.09	80.6	
O1151-03	EXGQ1	3	1.15	8.59	9.74	7.38	72.5	
O1151-04	EXGQ1MS	4	1.15	8.59	9.74	7.38	72.5	
O1151-05	EXGQ1MSD	5	1.15	8.59	9.74	7.38	72.5	
O1151-06	EXGQ2	6	1.15	8.67	9.82	7.45	72.7	
O1151-07	EXGQ3	7	1.16	8.70	9.86	8.21	81.0	
O1151-08	EXGQ4	8	1.15	8.79	9.94	8.5	83.6	
O1151-09	EXGQ5	9	1.12	8.85	9.97	8.25	80.6	
O1151-10	EXGQ6	10	1.13	8.53	9.66	7.89	79.2	
O1151-11	EXGQ7	11	1.15	8.81	9.96	8.57	84.2	
O1151-12	EXGQ8	12	1.15	8.81	9.96	8.53	83.8	
O1151-13	EXGQ9	13	1.13	8.77	9.9	8.68	86.1	
O1151-14	EXGR0	14	1.11	8.57	9.68	8.24	83.2	
O1151-15	EXGR1	15	1.13	8.53	9.66	8.37	84.9	
O1151-16	EXGR2	16	1.12	8.67	9.79	8.68	87.2	
O1151-17	EXGR3	17	1.13	8.64	9.77	8.65	87.0	
O1151-18	EXGR4	18	1.16	8.81	9.97	7.75	74.8	
O1151-19	EXGR5	19	1.16	8.60	9.76	8.15	81.3	
O1151-20	EXGR6	20	1.11	8.77	9.88	8.13	80.0	
O1151-21	EXGR7	21	1.19	8.70	9.89	7.87	76.8	
O1151-22	EXGR8	22	1.16	8.39	9.55	7.76	78.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$