



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB123689

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	01/13/2023	16:45
2	CAL2	1	Water	NA	NA	20.8	7.00	01/13/2023	16:46
3	CAL3	1	Water	NA	NA	20.8	10.03	01/13/2023	16:47
4	ICV	1	Water	NA	NA	20.8	6.99	01/13/2023	16:50
5	CCV1	1	Water	NA	NA	20.8	2.01	01/13/2023	16:52
6	O1155-01	1	Solid	20.02	20	20.9	6.88	01/13/2023	17:00
7	O1155-01DUP	1	Solid	20.03	20	20.8	6.89	01/13/2023	17:01
8	CCV2	1	Water	NA	NA	20.8	12.03	01/13/2023	17:05

1623689

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PHWO1134

WorkList ID : 166579

Department : Wet-Chemistry

Date : 01-13-2023 16:26:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/13/2023	Solid	O1155-01	pH		Cool 4 deg C	PSEG03 N51	34254	01/13/2023	9045D

Date/Time

01/13/23 16:35

Raw Sample Received by:

JD/WC1

Raw Sample Relinquished by:

RM WC1

Date/Time

01/13/23

Raw Sample Received by:

RM

Raw Sample Relinquished by:

JD/WC1