



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB123691

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	01/13/2023	16:45
2	CAL2	1	Water	NA	NA	20.8	7.00	01/13/2023	16:46
3	CAL3	1	Water	NA	NA	20.8	10.03	01/13/2023	16:47
4	ICV	1	Water	NA	NA	20.8	6.99	01/13/2023	16:50
5	CCV1	1	Water	NA	NA	20.8	2.01	01/13/2023	17:25
6	O1134-02	1	Water	NA	NA	20.6	8.05	01/13/2023	17:26
7	O1134-02DUP	1	Water	NA	NA	20.7	8.06	01/13/2023	17:27
8	CCV2	1	Water	NA	NA	20.8	12.03	01/13/2023	17:29

13123697

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH O1155 WorkList ID : 166578 Department : Wet-Chemistry Date : 01-13-2023 16:23:43

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/12/2023	Water	O1134-02	pH		Cool 4 deg C	PSEG03	N43	1115	01/12/2023 9040C

Date/Time 01.13.23 16:25
Raw Sample Received by: JO (WCL)
Raw Sample Relinquished by: JTCM7

Date/Time 01.13.23 18:00
Raw Sample Received by: JTCM7
Raw Sample Relinquished by: JO (WCL)