

LB123697


7471B

INSTRUMENT ID: CVL

Linear

 μ Abs.:

53908



A= 0.0000e+000

B= 1.8645e-004

C= -4.0551e-002

Rho= 0.9999521

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	U/D
0.0	0.000	0.007	0.007	254	0.000	254					-
0.2	0.200	0.191	-0.009	1242	0.0 %	1242					5
2.5	2.500	2.546	0.046	13871	0.0 %	13871					2
5.0	5.000	4.928	-0.072	26649	0.0 %	26649					-1
7.5	7.500	7.518	0.018	40541	0.0 %	40541					0
10.0	10.000	10.010	0.010	53908	0.0 %	53908					0

LB123697 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Jan 2023 10:12:49

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	254	-	0.0000	7471B	PPB	16 Jan 2023 10:23:20	S	Std
0.2 - 1	50.2	1242	-	0.2000	7471B	PPB	16 Jan 2023 10:25:37	S	Std
2.5 - 1	52.5	13871	-	2.5000	7471B	PPB	16 Jan 2023 10:27:54	S	Std
5.0 - 1	55	26649	-	5.0000	7471B	PPB	16 Jan 2023 10:30:11	S	Std
7.5 - 1	57.5	40541	-	7.5000	7471B	PPB	16 Jan 2023 10:32:28	S	Std
10.0 - 1	510	53908	-	10.0000	7471B	PPB	16 Jan 2023 10:34:45	S	Std
ICV02 - 1	ICV02	20413	3.7654	-	7471B	PPB	16 Jan 2023 10:37:29	U	SMPL
ICB02 - 1	ICB02	127	-0.0169	-	7471B	PPB	16 Jan 2023 10:39:44	U	SMPL
CCV55 - 1	CCV55	27041	5.0011	-	7471B	PPB	16 Jan 2023 10:41:59	U	SMPL
CCB55 - 1	CCB55	50	-0.0312	-	7471B	PPB	16 Jan 2023 10:44:14	U	SMPL
CRA - 1	CRA	1080	0.1608	-	7471B	PPB	16 Jan 2023 10:49:10	U	SMPL
HighStd - 1	HighStd	54141	10.0538	-	7471B	PPB	16 Jan 2023 10:51:25	U	SMPL
ChkStd - 1	ChkStd	38785	7.1907	-	7471B	PPB	16 Jan 2023 10:53:40	U	SMPL
PB150258BL - 1	PB	-9	-0.0422	-	7471B	PPB	16 Jan 2023 10:55:56	U	SMPL
PB150258BS - 1	LCSS	19455	3.5867	-	7471B	PPB	16 Jan 2023 10:58:12	U	SMPL
O1129-01 - 1	VNJ-207	59	-0.0296	-	7471B	PPB	16 Jan 2023 11:00:28	U	SMPL
O1132-01 - 1	COMPOSITE-1	7577	1.3721	-	7471B	PPB	16 Jan 2023 11:02:44	U	SMPL
O1133-01 - 1	COMPOSITE-1	7548	1.3667	-	7471B	PPB	16 Jan 2023 11:05:01	U	SMPL
O1133-03 - 1	COMPOSITE-2	9220	1.6785	-	7471B	PPB	16 Jan 2023 11:07:18	U	SMPL
O1135-01 - 1	WC-2	2447	0.4157	-	7471B	PPB	16 Jan 2023 11:09:33	U	SMPL
CCV56 - 1	CCV56	26979	4.9896	-	7471B	PPB	16 Jan 2023 11:11:48	U	SMPL
CCB56 - 1	CCB56	10	-0.0387	-	7471B	PPB	16 Jan 2023 11:14:03	U	SMPL
O1135-04 - 1	WC-3	18249	3.3619	-	7471B	PPB	16 Jan 2023 11:16:19	U	SMPL
O1140-01 - 1	TP-D	2690	0.4610	-	7471B	PPB	16 Jan 2023 11:18:35	U	SMPL
O1140-04 - 1	MH-5	396	0.0333	-	7471B	PPB	16 Jan 2023 11:20:50	U	SMPL
O1140-04DUP - 1	MH-5DUP	381	0.0305	-	7471B	PPB	16 Jan 2023 11:23:06	U	SMPL
O1140-04MS - 1	MH-5MS	18229	3.3582	-	7471B	PPB	16 Jan 2023 11:25:22	U	SMPL
O1140-04MSD - 1	MH-5MSD	20849	3.8467	-	7471B	PPB	16 Jan 2023 11:27:39	U	SMPL
O1140-04LX5 - 1		-67	-0.0530	-	7471B	PPB	16 Jan 2023 11:29:55	U	SMPL
O1140-04A - 1		20456	3.7734	-	7471B	PPB	16 Jan 2023 11:32:12	U	SMPL
O1100-05 - 1	WASTE	75667	14.0672	-	7471B	PPB	16 Jan 2023 11:34:29	U	SMPL
O1100-05DLX5 - 1		15654	2.8781	-	7471B	PPB	16 Jan 2023 11:41:48	U	SMPL
CCV57 - 1	CCV57	28207	5.2185	-	7471B	PPB	16 Jan 2023 11:44:03	U	SMPL
CCB57 - 1	CCB57	27	-0.0355	-	7471B	PPB	16 Jan 2023 11:46:19	U	SMPL