

LB129703


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43142



A= 0.0000e+000

B= 2.2954e-004

C= -1.1130e-002

Rho= 0.9997719

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0 D
0.0	0.000	-0.002	-0.002	41	0.000	41					-
0.2	0.200	0.195	-0.005	897	0.0 %	897					-3
2.5	2.500	2.421	-0.079	10596	0.0 %	10596					-3
5.0	5.000	5.085	0.085	22201	0.0 %	22201					2
7.5	7.500	7.610	0.110	33201	0.0 %	33201					1
10.0	10.000	9.891	-0.109	43142	0.0 %	43142					-1

LB123703 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 16 Jan 2023 12:39:27

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	41	-	0.0000	7470A	PPB	16 Jan 2023 12:40:58	S	Std
0.2 - 1	50.2	897	-	0.2000	7470A	PPB	16 Jan 2023 12:43:15	S	Std
2.5 - 1	52.5	10536	-	2.5000	7470A	PPB	16 Jan 2023 12:45:32	S	Std
5.0 - 1	55	22201	-	5.0000	7470A	PPB	16 Jan 2023 12:47:50	S	Std
7.5 - 1	57.5	33201	-	7.5000	7470A	PPB	16 Jan 2023 12:50:06	S	Std
10.0 - 1	57.0	43142	-	10.0000	7470A	PPB	16 Jan 2023 12:54:58	S	Std
ICV03 - 1	ICV03	16892	3.8662	-	7470A	PPB	16 Jan 2023 12:57:50	U	SMPL
ICB03 - 1	ICB03	64	0.0036	-	7470A	PPB	16 Jan 2023 13:00:05	U	SMPL
CCV58 - 1	CCV58	23129	5.2978	-	7470A	PPB	16 Jan 2023 13:02:21	U	SMPL
CCB58 - 1	CCB58	25	-0.0054	-	7470A	PPB	16 Jan 2023 13:04:38	U	SMPL
CRA - 1	CRA	861	0.1865	-	7470A	PPB	16 Jan 2023 13:06:51	U	SMPL
HighStd - 1	HighStd	43197	9.9041	-	7470A	PPB	16 Jan 2023 13:09:06	U	SMPL
ChkStd - 1	ChkStd	30967	7.0968	-	7470A	PPB	16 Jan 2023 13:11:21	U	SMPL
PB150271BL - 1	PBW	-40	-0.0203	-	7470A	PPB	16 Jan 2023 13:13:36	U	SMPL
PB150271BS - 1	LCSW	18135	4.1515	-	7470A	PPB	16 Jan 2023 13:15:52	U	SMPL
O1115-02 - 1	SAS-OIL	158	0.0251	-	7470A	PPB	16 Jan 2023 13:18:08	U	SMPL
O1115-04 - 1	SAS-OILY-DEBRIS	-11	-0.0137	-	7470A	PPB	16 Jan 2023 13:20:24	U	SMPL
O1124-02 - 1	TP-C	4	-0.0102	-	7470A	PPB	16 Jan 2023 13:22:41	U	SMPL
O1124-05 - 1	MH-4	-7	-0.0127	-	7470A	PPB	16 Jan 2023 13:24:58	U	SMPL
O1129-02 - 1	VNJ-207	-25	-0.0169	-	7470A	PPB	16 Jan 2023 13:27:14	U	SMPL
CCV59 - 1	CCV59	22979	5.2634	-	7470A	PPB	16 Jan 2023 13:29:29	U	SMPL
CCB59 - 1	CCB59	-18	-0.0153	-	7470A	PPB	16 Jan 2023 13:31:44	U	SMPL
O1132-02 - 1	COMPOSITE-1	-2	-0.0116	-	7470A	PPB	16 Jan 2023 13:33:59	U	SMPL
O1133-02 - 1	COMPOSITE-1	26	-0.0052	-	7470A	PPB	16 Jan 2023 13:36:15	U	SMPL
O1133-04 - 1	COMPOSITE-2	41	-0.0017	-	7470A	PPB	16 Jan 2023 13:38:31	U	SMPL
O1134-01 - 1	1114	29	-0.0045	-	7470A	PPB	16 Jan 2023 13:40:47	U	SMPL
O1135-02 - 1	WC-2	9	-0.0091	-	7470A	PPB	16 Jan 2023 13:43:03	U	SMPL
O1135-05 - 1	WC-3	1	-0.0109	-	7470A	PPB	16 Jan 2023 13:45:19	U	SMPL
O1140-02 - 1	TP-D	5	-0.0100	-	7470A	PPB	16 Jan 2023 13:47:36	U	SMPL
O1140-05 - 1	MH-5	9	-0.0091	-	7470A	PPB	16 Jan 2023 13:49:53	U	SMPL
O1140-05DUP - 1	MH-5DUP	32	-0.0038	-	7470A	PPB	16 Jan 2023 13:52:16	U	SMPL
O1140-05MS - 1	MH-5MS	18090	4.1412	-	7470A	PPB	16 Jan 2023 13:54:32	U	SMPL
CCV60 - 1	CCV60	22939	5.2542	-	7470A	PPB	16 Jan 2023 13:56:48	U	SMPL
CCB60 - 1	CCB60	-1	-0.0114	-	7470A	PPB	16 Jan 2023 13:59:04	U	SMPL
O1140-05MSD - 1	MH-5MSD	16677	3.8168	-	7470A	PPB	16 Jan 2023 14:01:19	U	SMPL
PB150272BL - 1	PBW	17	-0.0072	-	7470A	PPB	16 Jan 2023 14:03:35	U	SMPL
PB150272BS - 1	LCSW	18115	4.1469	-	7470A	PPB	16 Jan 2023 14:05:51	U	SMPL
O1159-02 - 1	TP-E	-7	-0.0127	-	7470A	PPB	16 Jan 2023 14:08:08	U	SMPL
O1159-05 - 1	MH-6	0	-0.0111	-	7470A	PPB	16 Jan 2023 14:10:22	U	SMPL
O1159-05DUP - 1	MH-6DUP	13	-0.0081	-	7470A	PPB	16 Jan 2023 14:12:38	U	SMPL
O1159-05MS - 1	MH-6MS	16161	3.6984	-	7470A	PPB	16 Jan 2023 14:14:55	U	SMPL
O1159-05MSD - 1	MH-6MSD	15782	3.6114	-	7470A	PPB	16 Jan 2023 14:17:12	U	SMPL
PB150288BL - 1	PBW	-35	-0.0192	-	7470A	PPB	16 Jan 2023 14:19:29	U	SMPL
PB150288BS - 1	LCSW	19009	4.3521	-	7470A	PPB	16 Jan 2023 14:21:44	U	SMPL
CCV61 - 1	CCV61	22529	5.1601	-	7470A	PPB	16 Jan 2023 14:24:00	U	SMPL
CCB61 - 1	CCB61	10	-0.0088	-	7470A	PPB	16 Jan 2023 14:26:16	U	SMPL
O1134-02 - 1	1115	69	0.0047	-	7470A	PPB	16 Jan 2023 14:28:32	U	SMPL
O1134-02DUP - 1	1115DUP	46	-0.0006	-	7470A	PPB	16 Jan 2023 14:30:48	U	SMPL
O1134-02MS - 1	1115MS	14292	3.2694	-	7470A	PPB	16 Jan 2023 14:33:04	U	SMPL
O1134-02MSD - 1	1115MSD	14525	3.3229	-	7470A	PPB	16 Jan 2023 14:35:19	U	SMPL
PB150273BL - 1	PBW	-2	-0.0116	-	7470A	PPB	16 Jan 2023 14:37:36	U	SMPL
PB150273BS - 1	LCSW	18116	4.1471	-	7470A	PPB	16 Jan 2023 14:39:52	U	SMPL
O1152-02 - 1	TOPSOIL-STMWTR-DENALI-2	8	-0.0093	-	7470A	PPB	16 Jan 2023 14:42:08	U	SMPL
O1152-02DUP - 1	TOPSOIL-STMWTR-DENALI-2DUP	29	-0.0045	-	7470A	PPB	16 Jan 2023 14:44:25	U	SMPL
O1152-02MS - 1	TOPSOIL-STMWTR-DENALI-2MS	14829	3.3926	-	7470A	PPB	16 Jan 2023 14:46:42	U	SMPL
O1152-02MSD - 1	TOPSOIL-STMWTR-DENALI-2MSD	15771	3.6089	-	7470A	PPB	16 Jan 2023 14:48:58	U	SMPL
CCV62 - 1	CCV62	21654	4.9592	-	7470A	PPB	16 Jan 2023 14:51:14	U	SMPL
CCB62 - 1	CCB62	0	-0.0111	-	7470A	PPB	16 Jan 2023 14:53:30	U	SMPL
N6245-02 - 1	WASTE-COMP	34	-0.0033	-	7470A	PPB	16 Jan 2023 14:55:46	U	SMPL
PB150222TB - 1	PB150222TB	37	-0.0026	-	7470A	PPB	16 Jan 2023 14:58:02	U	SMPL
PB150253TB - 1	PB150253TB	7	-0.0095	-	7470A	PPB	16 Jan 2023 15:00:18	U	SMPL
PB150254TB - 1	PB150254TB	-31	-0.0182	-	7470A	PPB	16 Jan 2023 15:02:33	U	SMPL
PB150285TB - 1	PB150285TB	-14	-0.0143	-	7470A	PPB	16 Jan 2023 15:04:50	U	SMPL
O1140-05LX5 - 1		-40	-0.0203	-	7470A	PPB	16 Jan 2023 15:07:06	U	SMPL
O1140-05A - 1		16367	3.7457	-	7470A	PPB	16 Jan 2023 15:09:23	U	SMPL
O1159-05LX5 - 1		-58	-0.0240	-	7470A	PPB	16 Jan 2023 15:11:40	U	SMPL
O1159-05A - 1		16467	3.7686	-	7470A	PPB	16 Jan 2023 15:13:57	U	SMPL
CCV63 - 1	CCV63	20732	4.7476	-	7470A	PPB	16 Jan 2023 15:16:13	U	SMPL
CCB63 - 1	CCB63	-14	-0.0143	-	7470A	PPB	16 Jan 2023 15:18:30	U	SMPL
O1134-02LX5 - 1		-17	-0.0150	-	7470A	PPB	16 Jan 2023 15:20:46	U	SMPL
O1134-02A - 1		16281	3.7259	-	7470A	PPB	16 Jan 2023 15:25:55	U	SMPL
O1152-02LX5 - 1		46	-0.0006	-	7470A	PPB	16 Jan 2023 15:28:11	U	SMPL
O1152-02A - 1		17231	3.9440	-	7470A	PPB	16 Jan 2023 15:30:27	U	SMPL
CCV64 - 1	CCV64	22200	5.0846	-	7470A	PPB	16 Jan 2023 15:32:43	U	SMPL
CCB64 - 1	CCB64	-14	-0.0143	-	7470A	PPB	16 Jan 2023 15:34:59	U	SMPL