

LB123712


7471B

INSTRUMENT ID: EV4

Linear

 μ Abs.:

51913



A= 0.0000e+000

B= 1.9188e-004

C= 5.2127e-002

Rho= 0.9998588

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.049	0.049	-15	0.000	-15					0/0
0.2	0.200	0.234	0.034	949	0.0 %	949					17
2.5	2.500	2.367	-0.133	12065	0.0 %	12065					5
5.0	5.000	5.041	0.041	25997	0.0 %	25997					1
7.5	7.500	7.495	-0.005	38791	0.0 %	38791					0
10.0	10.000	10.013	0.013	51913	0.0 %	51913					0

LB123712 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 17 Jan 2023 10:40:46

Sample ID	Extended ID	μ Abs	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	-15	-	0.00007471B	PPB	17 Jan 2023 10:46:21	S	Std
0.2 - 1	50.2	949	-	0.20007471B	PPB	17 Jan 2023 10:48:38	S	Std
2.5 - 1	52.5	12065	-	2.50007471B	PPB	17 Jan 2023 10:50:55	S	Std
5.0 - 1	55	25997	-	5.00007471B	PPB	17 Jan 2023 10:53:12	S	Std
7.5 - 1	57.5	38791	-	7.50007471B	PPB	17 Jan 2023 10:55:28	S	Std
10.0 - 1	51.1	51913	-	10.00007471B	PPB	17 Jan 2023 10:57:44	S	Std
ICV04 - 1	ICV04	20202	3.9285	-7471B	PPB	17 Jan 2023 11:00:43	U	SMPL
ICB04 - 1	ICB04	5	0.0531	-7471B	PPB	17 Jan 2023 11:02:58	U	SMPL
CCV63 - 1	CCV63	26106	5.0614	-7471B	PPB	17 Jan 2023 11:05:13	U	SMPL
CCB63 - 1	CCB63	-2	0.0517	-7471B	PPB	17 Jan 2023 11:07:28	U	SMPL
CRA - 1	CRA	961	0.2365	-7471B	PPB	17 Jan 2023 11:09:43	U	SMPL
HighStd - 1	HighStd	52752	10.1743	-7471B	PPB	17 Jan 2023 11:11:58	U	SMPL
ChkStd - 1	ChkStd	37049	7.1612	-7471B	PPB	17 Jan 2023 11:14:14	U	SMPL
PB150305BL - 1	PBS	-74	0.0379	-7471B	PPB	17 Jan 2023 11:16:30	U	SMPL
PB150305BS - 1	LCSS	20188	3.9259	-7471B	PPB	17 Jan 2023 11:22:16	U	SMPL
O1141-01 - 1	WEST-WALL	303	0.1103	-7471B	PPB	17 Jan 2023 11:24:31	U	SMPL
O1141-02 - 1	SOUTH-WALL	1793	0.3962	-7471B	PPB	17 Jan 2023 11:26:48	U	SMPL
O1141-03 - 1	EAST-WALL	781	0.2020	-7471B	PPB	17 Jan 2023 11:29:05	U	SMPL
O1152-01 - 1	TOPSOIL-STMWTR-DENALI-2	1211	0.2845	-7471B	PPB	17 Jan 2023 11:31:22	U	SMPL
O1155-01 - 1	34254	6	0.0533	-7471B	PPB	17 Jan 2023 11:33:37	U	SMPL
CCV64 - 1	CCV64	26076	5.0557	-7471B	PPB	17 Jan 2023 11:35:53	U	SMPL
CCB64 - 1	CCB64	-73	0.0381	-7471B	PPB	17 Jan 2023 11:38:08	U	SMPL
O1157-01 - 1	TR-4-011323	2092	0.4535	-7471B	PPB	17 Jan 2023 11:40:23	U	SMPL
O1159-01 - 1	TP-E	2802	0.5898	-7471B	PPB	17 Jan 2023 11:42:38	U	SMPL
O1159-04 - 1	MH-6	343	0.1179	-7471B	PPB	17 Jan 2023 11:44:54	U	SMPL
O1160-01 - 1	B-04-SB02	1303	0.3022	-7471B	PPB	17 Jan 2023 11:47:10	U	SMPL
O1161-01 - 1	20230113-H4-STOCKPILE	34334	6.6402	-7471B	PPB	17 Jan 2023 11:49:25	U	SMPL
O1167-01 - 1	EO-3-011623	2432	0.5188	-7471B	PPB	17 Jan 2023 11:51:41	U	SMPL
O1170-01 - 1	CL-2-011623	2448	0.5219	-7471B	PPB	17 Jan 2023 11:58:47	U	SMPL
O1170-01DUP - 1	CL-2-011623DUP	2076	0.4505	-7471B	PPB	17 Jan 2023 12:01:03	U	SMPL
O1170-01MS - 1	CL-2-011623MS	24847	4.8198	-7471B	PPB	17 Jan 2023 12:03:21	U	SMPL
O1170-01MSD - 1	CL-2-011623MSD	24208	4.6972	-7471B	PPB	17 Jan 2023 12:05:37	U	SMPL
CCV65 - 1	CCV65	25456	4.9367	-7471B	PPB	17 Jan 2023 12:07:52	U	SMPL
CCB65 - 1	CCB65	-145	0.0243	-7471B	PPB	17 Jan 2023 12:10:08	U	SMPL
O1172-01 - 1	TP-F	15528	3.0317	-7471B	PPB	17 Jan 2023 12:12:23	U	SMPL
O1170-01LX5 - 1		267	0.1034	-7471B	PPB	17 Jan 2023 12:14:39	U	SMPL
O1170-01A - 1		24087	4.6740	-7471B	PPB	17 Jan 2023 12:16:54	U	SMPL
CCV66 - 1	CCV66	25458	4.9371	-7471B	PPB	17 Jan 2023 12:19:10	U	SMPL
CCB66 - 1	CCB66	-57	0.0412	-7471B	PPB	17 Jan 2023 12:21:26	U	SMPL