

L8123716


7470A

INSTRUMENT ID: CV2

Linear

μ Abs.:

47115



A= 0.0000e+000

B= 2.1253e-004

C= 3.4427e-002

Rho= 0.9977632

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.0	0.000	0.039	0.039	22	0.000	22					-1
0.2	0.200	0.199	-0.001	774	0.0 %	774					-2
2.5	2.500	2.315	-0.185	10730	0.0 %	10730					9
5.0	5.000	5.451	0.451	25485	0.0 %	25485					-5
7.5	7.500	7.148	-0.352	33472	0.0 %	33472					0
10.0	10.000	10.048	0.048	47119	0.0 %	47119					

LB123716 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Jan 2023 13:42:31

Sample ID	Extended ID	Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	56	22	-	0.0000	7470A	PPB	17 Jan 2023 13:48:06	S	Std
0.2 - 1	50-2	774	-	0.2000	7470A	PPB	17 Jan 2023 13:50:23	S	Std
2.5 - 1	52-5	10730	-	2.5000	7470A	PPB	17 Jan 2023 13:52:40	S	Std
5.0 - 1	55	25485	-	5.0000	7470A	PPB	17 Jan 2023 13:54:57	S	Std
7.5 - 1	57-5	33472	-	7.5000	7470A	PPB	17 Jan 2023 13:57:14	S	Std
10.0 - 1	510	47119	-	10.0000	7470A	PPB	17 Jan 2023 14:02:20	S	Std
ICV05 - 1	ICV05	16966	3.6401	-	7470A	PPB	17 Jan 2023 14:05:27	U	SMPL
ICB05 - 1	ICB05	34	0.0417	-	7470A	PPB	17 Jan 2023 14:07:41	U	SMPL
CCV67 - 1	CCV67	24541	5.2500	-	7470A	PPB	17 Jan 2023 14:09:56	U	SMPL
CCB67 - 1	CCB67	-14	0.0315	-	7470A	PPB	17 Jan 2023 14:12:11	U	SMPL
CRA - 1	CRA	787	0.2017	-	7470A	PPB	17 Jan 2023 14:14:26	U	SMPL
HighStd - 1	HighStd	46694	9.9581	-	7470A	PPB	17 Jan 2023 14:16:41	U	SMPL
ChkStd - 1	ChkStd	31096	6.6431	-	7470A	PPB	17 Jan 2023 14:18:56	U	SMPL
PB150308BL - 1	PBW	-13	0.0317	-	7470A	PPB	17 Jan 2023 14:21:12	U	SMPL
PB150308BS - 1	LCSW	15950	3.4242	-	7470A	PPB	17 Jan 2023 14:23:29	U	SMPL
O1092-01 - 1	FMI-120	81	0.0516	-	7470A	PPB	17 Jan 2023 14:25:45	U	SMPL
O1092-01DUP - 1	FMI-120DUP	84	0.0523	-	7470A	PPB	17 Jan 2023 14:27:01	U	SMPL
O1092-01MS - 1	FMI-120MS	13570	2.9184	-	7470A	PPB	17 Jan 2023 14:29:18	U	SMPL
O1092-01MSD - 1	FMI-120MSD	14399	3.0946	-	7470A	PPB	17 Jan 2023 14:31:35	U	SMPL
O1092-03 - 1	RL-TOTE-0110	496608	HIGH	-	7470A	PPB	17 Jan 2023 14:33:50	X	
CCV68 - 1	CCV68	23771	5.0864	-	7470A	PPB	17 Jan 2023 14:35:38	U	SMPL
CCB68 - 1	CCB68	-109	0.0113	-	7470A	PPB	17 Jan 2023 14:37:56	U	SMPL
O1125-02 - 1	EB-2023-1-11	109	0.0576	-	7470A	PPB	17 Jan 2023 14:40:11	U	SMPL
O1160-02 - 1	EB-2023-1-13	47	0.0444	-	7470A	PPB	17 Jan 2023 14:42:26	U	SMPL
O1092-03DLX10 - 1		74022	15.7660	-	7470A	PPB	17 Jan 2023 14:44:42	U	SMPL
O1092-01LX5 - 1		-308	-0.0310	-	7470A	PPB	17 Jan 2023 14:46:57	U	SMPL
O1092-01A - 1		16675	3.5783	-	7470A	PPB	17 Jan 2023 14:49:15	U	SMPL
O1092-03DL2X50 - 1		14811	3.1821	-	7470A	PPB	17 Jan 2023 14:51:31	U	SMPL
CCV69 - 1	CCV69	23871	5.1076	-	7470A	PPB	17 Jan 2023 14:53:48	U	SMPL
CCB69 - 1	CCB69	-138	0.0051	-	7470A	PPB	17 Jan 2023 14:56:05	U	SMPL