



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB123722

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	01/17/2023	16:35
2	CAL2	1	Water	NA	NA	20.8	7.00	01/17/2023	16:36
3	CAL3	1	Water	NA	NA	20.7	10.03	01/17/2023	16:37
4	ICV	1	Water	NA	NA	20.8	6.99	01/17/2023	16:40
5	CCV1	1	Water	NA	NA	20.7	2.01	01/17/2023	16:42
6	O1172-02	1	Solid	20.02	20	20.9	8.73	01/17/2023	16:45
7	O1180-02	1	Solid	20.01	20	20.2	7.57	01/17/2023	16:48
8	O1180-02DUP	1	Solid	20.02	20	20.3	7.58	01/17/2023	16:49
9	CCV2	1	Water	NA	NA	20.8	12.03	01/17/2023	16:50

123422

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIIVITY O1180

WorkList ID : 166646

Department : Wet-Chemistry

Date : 01-17-2023 13:59:00

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/16/2023	Solid	O1172-02	Corrosivity	Cool 4 deg C	PSEG03	N61	TP-F	01/16/2023	9045D
01/17/2023	Solid	O1180-02	Corrosivity	Cool 4 deg C	PSEG03	N11	MH-7	01/17/2023	9045D

Date/Time 01-17-23 16:10
Raw Sample Received by: 706201
Raw Sample Relinquished by: 706201

Date/Time 01-17-23 17:20
Raw Sample Received by: 706201
Raw Sample Relinquished by: 706201