

LB123726


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

4498E



A= 0.0000e+000

B= 2.2570e-004 *slp*C= -1.2533e-001 *intercept*

Rho= 0.9998783

Accept= Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>%D</i>
0.00	0.000	0.041	0.041	739	0.000	739					<i>11</i>
0.05	0.050					185	0				<i>1</i>
0.20	0.200	0.222	0.022	1540	0.0 %	1540					<i>2</i>
2.50	2.500	2.484	-0.016	11559	0.0 %	11559					<i>1</i>
5.00	5.000	4.879	-0.121	22174	0.0 %	22174					<i>0</i>
7.50	7.500	7.546	0.046	33987	0.0 %	33987					
10.0	10.000	10.028	0.028	44985	0.0 %	44985					

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	739	-		0 SFAM01.1	PPB	1/18/2023 13:44	Std
	0.2 S01	1540	-		0.2 SFAM01.1	PPB	1/18/2023 13:46	Std
	2.5 S02	11559	-		2.5 SFAM01.1	PPB	1/18/2023 13:49	Std
	5 S03	22174	-		5 SFAM01.1	PPB	1/18/2023 13:51	Std
	7.5 S04	33987	-		7.5 SFAM01.1	PPB	1/18/2023 13:53	Std
	10 S05	44985	-		10 SFAM01.1	PPB	1/18/2023 13:58	Std
ICV006	ICV006	17655	3.8594 -		SFAM01.1	PPB	1/18/2023 14:01	SMPL
ICB006	ICB006	929	0.0843 -		SFAM01.1	PPB	1/18/2023 14:04	SMPL
CCV070	CCV070	22282	4.9037 -		SFAM01.1	PPB	1/18/2023 14:06	SMPL
CCB070	CCB070	889	0.0753 -		SFAM01.1	PPB	1/18/2023 14:08	SMPL
PB150320BL	PBW320	488	-0.0152 -		SFAM01.1	PPB	1/18/2023 14:13	SMPL
O1116-01	YBKK0	3069	0.5673 -		SFAM01.1	PPB	1/18/2023 14:15	SMPL
O1116-02	YBKK2	9790	2.0843 -		SFAM01.1	PPB	1/18/2023 14:17	SMPL
O1116-03	YBKK4	38658	8.5998 -		SFAM01.1	PPB	1/18/2023 14:20	SMPL
O1116-04	YBKK6	35632	7.9168 -		SFAM01.1	PPB	1/18/2023 14:22	SMPL
O1116-05	YBKLO	20539	4.5103 -		SFAM01.1	PPB	1/18/2023 14:24	SMPL
O1116-06	YBKL2	1092	0.1211 -		SFAM01.1	PPB	1/18/2023 14:27	SMPL
O1116-07	YBKL4	1125	0.1286 -		SFAM01.1	PPB	1/18/2023 14:29	SMPL
O1116-08	YBKL6	1482	0.2092 -		SFAM01.1	PPB	1/18/2023 14:31	SMPL
O1116-09	YBKL8	1911	0.306 -		SFAM01.1	PPB	1/18/2023 14:33	SMPL
O1116-10	YBKMO	2047	0.3367 -		SFAM01.1	PPB	1/18/2023 14:36	SMPL
O1116-11	YBKMO8	59800	13.3716 -		SFAM01.1	PPB	1/18/2023 14:38	SMPL
O1116-12	YBKNO	1652	0.2475 -		SFAM01.1	PPB	1/18/2023 14:40	SMPL
O1116-13	YBKN2	943	0.0875 -		SFAM01.1	PPB	1/18/2023 14:42	SMPL
O1116-14	YBKN2D	907	0.0794 -		SFAM01.1	PPB	1/18/2023 14:45	SMPL
O1116-15	YBKN2S	5105	1.0269 -		SFAM01.1	PPB	1/18/2023 14:47	SMPL
O1116-16	YBKN4	2835	0.5145 -		SFAM01.1	PPB	1/18/2023 14:49	SMPL
O1116-17	YBKN6	2977	0.5466 -		SFAM01.1	PPB	1/18/2023 14:51	SMPL
O1116-18	YBKN8	2845	0.5168 -		SFAM01.1	PPB	1/18/2023 14:54	SMPL
O1116-19	YBKP2	979	0.0956 -		SFAM01.1	PPB	1/18/2023 14:56	SMPL
CCV071	CCV071	22070	4.8559 -		SFAM01.1	PPB	1/18/2023 14:58	SMPL
CCB071	CCB071	852	0.067 -		SFAM01.1	PPB	1/18/2023 15:01	SMPL
O1116-20	YBKP4	1452	0.2024 -		SFAM01.1	PPB	1/18/2023 15:03	SMPL
O1116-21	YBKP8	17484	3.8208 -		SFAM01.1	PPB	1/18/2023 15:05	SMPL
O1116-22	YBKQ2	12774	2.7578 -		SFAM01.1	PPB	1/18/2023 15:07	SMPL
PB150321BL	PBW321	597	0.0094 -		SFAM01.1	PPB	1/18/2023 15:10	SMPL
O1117-01	YBKM4	5330	1.0777 -		SFAM01.1	PPB	1/18/2023 15:12	SMPL
O1117-02	YBKM6	1295	0.1669 -		SFAM01.1	PPB	1/18/2023 15:14	SMPL
O1117-03	YBKQ4	400	-0.0351 -		SFAM01.1	PPB	1/18/2023 15:16	SMPL
O1117-04	YBKP6	4409	0.8698 -		SFAM01.1	PPB	1/18/2023 15:19	SMPL
O1117-05	YBKP6D	3982	0.7734 -		SFAM01.1	PPB	1/18/2023 15:21	SMPL
O1117-06	YBKP6S	8282	1.7439 -		SFAM01.1	PPB	1/18/2023 15:23	SMPL
O1117-07	YBKQ6	216	-0.0766 -		SFAM01.1	PPB	1/18/2023 15:25	SMPL
O1117-08	YBKQ8	5512	1.1187 -		SFAM01.1	PPB	1/18/2023 15:28	SMPL
O1117-09	YBKRO	1761	0.2721 -		SFAM01.1	PPB	1/18/2023 15:30	SMPL
O1117-10	YBKR2	11162	2.3939 -		SFAM01.1	PPB	1/18/2023 15:32	SMPL

LB123726

INSTRUMENT ID : CV1

O1117-11	YBKR4	1062	0.1144 -	SFAM01.1	PPB	1/18/2023 15:35 SMPL
O1117-12	YBKR6	1077	0.1177 -	SFAM01.1	PPB	1/18/2023 15:37 SMPL
O1117-13	YBKR8	878	0.0728 -	SFAM01.1	PPB	1/18/2023 15:39 SMPL
O1117-14	YBKS0	1980	0.3216 -	SFAM01.1	PPB	1/18/2023 15:41 SMPL
O1118-01	YBKK1	633	0.0175 -	SFAM01.1	PPB	1/18/2023 15:46 SMPL
CCV072	CCV072	21692	4.7706 -	SFAM01.1	PPB	1/18/2023 15:48 SMPL
CCB072	CCB072	654	0.0223 -	SFAM01.1	PPB	1/18/2023 15:50 SMPL
O1118-02	YBKK3	654	0.0223 -	SFAM01.1	PPB	1/18/2023 15:53 SMPL
O1118-03	YBKK5	1143	0.1326 -	SFAM01.1	PPB	1/18/2023 15:55 SMPL
O1118-16	YBKN5	840	0.0643 -	SFAM01.1	PPB	1/18/2023 16:25 SMPL
O1118-17	YBKN7	54	-0.1131 -	SFAM01.1	PPB	1/18/2023 16:27 SMPL
O1118-18	YBKN9	118	-0.0987 -	SFAM01.1	PPB	1/18/2023 16:29 SMPL
O1118-19	YBKP3	208	-0.0784 -	SFAM01.1	PPB	1/18/2023 16:32 SMPL
O1118-20	YBKP5	564	0.002 -	SFAM01.1	PPB	1/18/2023 16:34 SMPL
O1118-21	YBKP9	161	-0.089 -	SFAM01.1	PPB	1/18/2023 16:36 SMPL
CCV073	CCV073	19721	4.3257 -	SFAM01.1	PPB	1/18/2023 16:38 SMPL
CCB073	CCB073	636	0.0182 -	SFAM01.1	PPB	1/18/2023 16:41 SMPL
O1118-22	YBKQ3	116	-0.0992 -	SFAM01.1	PPB	1/18/2023 16:43 SMPL
PB150323BL	PBW323	638	0.0187 -	SFAM01.1	PPB	1/18/2023 16:45 SMPL
O1119-01	YBKM5	398	-0.0355 -	SFAM01.1	PPB	1/18/2023 16:47 SMPL
O1119-02	YBKM7	217	-0.0764 -	SFAM01.1	PPB	1/18/2023 16:50 SMPL
O1119-03	YBKQ5	95	-0.1039 -	SFAM01.1	PPB	1/18/2023 16:52 SMPL
O1119-04	YBKP7	142	-0.0933 -	SFAM01.1	PPB	1/18/2023 16:54 SMPL
O1119-05	YBKP7D	150	-0.0915 -	SFAM01.1	PPB	1/18/2023 16:57 SMPL
O1119-06	YBKP7S	4413	0.8707 -	SFAM01.1	PPB	1/18/2023 16:59 SMPL
O1119-07	YBKQ7	74	-0.1086 -	SFAM01.1	PPB	1/18/2023 17:01 SMPL
O1119-08	YBKQ9	382	-0.0391 -	SFAM01.1	PPB	1/18/2023 17:03 SMPL
O1119-09	YBKR1	720	0.0372 -	SFAM01.1	PPB	1/18/2023 17:06 SMPL
O1119-10	YBKR3	806	0.0566 -	SFAM01.1	PPB	1/18/2023 17:08 SMPL
O1119-11	YBKR5	727	0.0388 -	SFAM01.1	PPB	1/18/2023 17:10 SMPL
O1119-12	YBKR7	695	0.0315 -	SFAM01.1	PPB	1/18/2023 17:12 SMPL
O1119-13	YBKR9	730	0.0394 -	SFAM01.1	PPB	1/18/2023 17:15 SMPL
O1119-14	YBKS1	878	0.0728 -	SFAM01.1	PPB	1/18/2023 17:17 SMPL
O1119-15	YBKS2	695	0.0315 -	SFAM01.1	PPB	1/18/2023 17:19 SMPL
O1119-16	YBKT9	724	0.0381 -	SFAM01.1	PPB	1/18/2023 17:22 SMPL
PB150322BL	PBW322	655	0.0225 -	SFAM01.1	PPB	1/18/2023 17:24 SMPL
CCV074	CCV074	19859	4.3569 -	SFAM01.1	PPB	1/18/2023 17:28 SMPL
CCB074	CCB074	631	0.0171 -	SFAM01.1	PPB	1/18/2023 17:31 SMPL
O1118-04		1064	0.1148 -	SFAM01.1	PPB	1/18/2023 17:33 SMPL
O1118-05		850	0.0665 -	SFAM01.1	PPB	1/18/2023 17:35 SMPL
O1118-06		431	-0.0281 -	SFAM01.1	PPB	1/18/2023 17:37 SMPL
O1118-07		460	-0.0215 -	SFAM01.1	PPB	1/18/2023 17:40 SMPL
O1118-08		411	-0.0326 -	SFAM01.1	PPB	1/18/2023 17:42 SMPL
O1118-09		257	-0.0673 -	SFAM01.1	PPB	1/18/2023 17:44 SMPL
O1118-10		472	-0.0188 -	SFAM01.1	PPB	1/18/2023 17:47 SMPL
O1118-11		761	0.0464 -	SFAM01.1	PPB	1/18/2023 17:49 SMPL
O1118-12		127	-0.0967 -	SFAM01.1	PPB	1/18/2023 17:51 SMPL

LB123726

INSTRUMENT ID : CV1

O1118-13		87	-0.1057 -	SFAM01.1	PPB	1/18/2023 17:53 SMPL
O1118-14		132	-0.0955 -	SFAM01.1	PPB	1/18/2023 17:56 SMPL
O1118-15		4421	0.8725 -	SFAM01.1	PPB	1/18/2023 17:58 SMPL
CCV075	CCV075	19984	4.3851 -	SFAM01.1	PPB	1/18/2023 18:00 SMPL
CCB075	CCB075	606	0.0114 -	SFAM01.1	PPB	1/18/2023 18:03 SMPL