



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123733

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	01/19/2023	08:30
2	O1199-01		1	12	100.00	>212.0	>212.0	01/19/2023	09:00
3	O1208-01		1	12	100.00	>212.0	>212.0	01/19/2023	09:30
4	O1208-01DUP		1	14	100.00	>212.0	>212.0	01/19/2023	10:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6123733

WorkList Name : fp1-17

WorkList ID : 166627

Department : Wet-Chemistry

Date : 01-17-2023 08:25:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/23/2023	Water	O1199-01	Flash Point	Cool 4 deg C	PSEG03	N62	34185	01/18/2023	1010B
01/25/2023	Water	O1208-01	Flash Point	Cool 4 deg C	PSEG03	N61	33894	01/18/2023	1010B

Date/Time 01/19/2023 08:20
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: JH(CWC)

Date/Time 01/19/2023 10:40
 Raw Sample Received by: JH(CWC)
 Raw Sample Relinquished by: JH CWC