



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB123734

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	01/19/2023	07:05
2	CAL2	1	Water	NA	NA	20.7	7.00	01/19/2023	07:06
3	CAL3	1	Water	NA	NA	20.8	10.03	01/19/2023	07:08
4	ICV	1	Water	NA	NA	20.7	6.99	01/19/2023	07:10
5	CCV1	1	Water	NA	NA	20.7	2.01	01/19/2023	07:11
6	O1192-01	1	Water	NA	NA	20.2	4.45	01/19/2023	07:15
7	O1192-02	1	Water	NA	NA	20.3	4.46	01/19/2023	07:17
8	O1192-03	1	Water	NA	NA	20.5	5.73	01/19/2023	07:19
9	O1199-01	1	Water	NA	NA	22.6	8.77	01/19/2023	07:26
10	O1208-01	1	Water	NA	NA	22.7	6.54	01/19/2023	07:30
11	O1208-01DUP	1	Water	NA	NA	22.8	6.55	01/19/2023	07:31
12	CCV2	1	Water	NA	NA	20.8	12.03	01/19/2023	07:33

UB 123134

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o1208 w WorkList ID : 166677 Department : Wet-Chemistry Date : 01-19-2023 05:52:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/09/2023	Water	O1192-01	pH	Cool 4 deg C	STER09	N11	CELL-A-POOL-WATER	01/09/2023	9040C
01/09/2023	Water	O1192-02	pH	Cool 4 deg C	STER09	N11	CELL-B/C-POOL-WATER	01/09/2023	9040C
01/09/2023	Water	O1192-03	pH	Cool 4 deg C	STER09	N11	CELL-D-POOL-WATER	01/09/2023	9040C
01/18/2023	Water	O1199-01	pH	Cool 4 deg C	PSEG03	N62	34185	01/18/2023	9040C
01/18/2023	Water	O1208-01	pH	Cool 4 deg C	PSEG03	N61	33894	01/18/2023	9040C

Date/Time 01-19-23 07:00
Raw Sample Received by: JPC
Raw Sample Relinquished by: JPC

Date/Time 01-19-23 10:00
Raw Sample Received by: JPC
Raw Sample Relinquished by: JPC