



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB123735

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	01/19/2023	07:05
2	CAL2	1	Water	NA	NA	20.7	7.00	01/19/2023	07:06
3	CAL3	1	Water	NA	NA	20.8	10.03	01/19/2023	07:08
4	ICV	1	Water	NA	NA	20.7	6.99	01/19/2023	07:10
5	CCV1	1	Water	NA	NA	20.8	2.01	01/19/2023	07:37
6	O1178-01	1	Solid	20.02	20	20.6	7.47	01/19/2023	07:40
7	O1215-01	1	Solid	20.03	20	20.7	7.68	01/19/2023	07:44
8	O1215-01DUP	1	Solid	20.03	20	20.7	7.69	01/19/2023	07:45
9	CCV2	1	Water	NA	NA	20.8	12.03	01/19/2023	07:47

VB 123435

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph s o1198

WorkList ID : 166678

Department : Wet-Chemistry

Date : 01-19-2023 05:53:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/17/2023	Solid	O1178-01	pH			A	Cool 4 deg C PSEG05 N61 AU-5-011723	01/17/2023	9045D
01/18/2023	Solid	O1215-01	pH			A	Cool 4 deg C PSEG05 N62 BU-02-011823	01/18/2023	9045D

Date/Time

01-19-23 07:00

Raw Sample Received by:

JD (WC)

Raw Sample Relinquished by:

CP SM

Date/Time

01-19-23

Raw Sample Received by:

CP SM

Raw Sample Relinquished by:

JD (WC)