



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB123736

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	01/19/2023	07:05
2	CAL2	1	Water	NA	NA	20.7	7.00	01/19/2023	07:06
3	CAL3	1	Water	NA	NA	20.8	10.03	01/19/2023	07:08
4	ICV	1	Water	NA	NA	20.7	6.99	01/19/2023	07:10
5	CCV1	1	Water	NA	NA	20.6	2.01	01/19/2023	08:00
6	O1214-03	1	Solid	20.02	20	21.3	7.83	01/19/2023	08:05
7	O1214-06	1	Solid	20.01	20	21.6	7.87	01/19/2023	08:06
8	O1214-09	1	Solid	20.02	20	21.5	7.92	01/19/2023	08:10
9	O1214-12	1	Solid	20.03	20	21.6	8.36	01/19/2023	08:12
10	O1214-15	1	Solid	20.04	20	21.5	8.25	01/19/2023	08:15
11	O1217-02	1	Solid	20.02	20	20.4	7.46	01/19/2023	08:20
12	O1217-15	1	Solid	20.03	20	20.8	7.33	01/19/2023	08:23
13	O1220-02	1	Solid	20.02	20	20.8	8.24	01/19/2023	08:25
14	O1220-05	1	Solid	20.03	20	20.7	8.25	01/19/2023	08:26
15	O1220-05DUP	1	Solid	20.04	20	20.8	8.26	01/19/2023	08:27
16	CCV2	1	Water	NA	NA	20.7	12.03	01/19/2023	08:30

123736

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY O1220 WorkList ID : 166691 Department : Wet-Chemistry Date : 01-19-2023 08:39:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/17/2023	Solid	O1214-03	Corrosivity	Cool 4 deg C	ENTE01	N62	TJRC-011723-B3-0-10	01/17/2023	9045D
01/17/2023	Solid	O1214-06	Corrosivity	Cool 4 deg C	ENTE01	N62	TJRC-011723-B3-10-20	01/17/2023	9045D
01/17/2023	Solid	O1214-09	Corrosivity	Cool 4 deg C	ENTE01	N62	TJRC-011723-B3-0-10-DUP	01/17/2023	9045D
01/18/2023	Solid	O1214-12	Corrosivity	Cool 4 deg C	ENTE01	N62	TJRC-011823-B3-20-30	01/18/2023	9045D
01/18/2023	Solid	O1214-15	Corrosivity	Cool 4 deg C	ENTE01	N62	TJRC-011823-B3-30-40	01/18/2023	9045D
01/18/2023	Solid	O1217-02	Corrosivity	Cool 4 deg C	PSEG03	N62	BORING-4	01/18/2023	9045D
01/18/2023	Solid	O1217-15	Corrosivity	Cool 4 deg C	PSEG03	N62	BORING-2	01/18/2023	9045D
01/18/2023	Solid	O1220-02	Corrosivity	Cool 4 deg C	PSEG03	N62	MH-8	01/18/2023	9045D
01/18/2023	Solid	O1220-05	Corrosivity	Cool 4 deg C	PSEG03	N62	TP-G	01/18/2023	9045D

Date/Time 01.19.23 08:40
 Raw Sample Received by: JP (CWC)
 Raw Sample Relinquished by: CP SM

Date/Time 01.19.23 10:40
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP (CWC)