

CB122739


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

43815



A= 0.0000e+000

B= 2.3227e-004 slope

C= -1.4749e-001 y-intercept

Rho= 0.9998264

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.076	0.076	962	0.000	962					=
0.05	0.050					185	0				2
0.20	0.200	0.203	0.003	1511	0.0 %	1511					-1
2.50	2.500	2.468	-0.032	11259	0.0 %	11259					-3
5.00	5.000	4.868	-0.132	21591	0.0 %	21591					-1
7.50	7.500	7.556	0.056	33165	0.0 %	33165					0
10.0	10.000	10.030	0.030	43815	0.0 %	43815					

LB123739

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	962	-		0 SFAM01.1	PPB	1/19/2023 10:51	Std
	0.2 S01	1511	-		0.2 SFAM01.1	PPB	1/19/2023 10:53	Std
	2.5 S02	11259	-		2.5 SFAM01.1	PPB	1/19/2023 10:56	Std
	5 S03	21591	-		5 SFAM01.1	PPB	1/19/2023 10:58	Std
	7.5 S04	33165	-		7.5 SFAM01.1	PPB	1/19/2023 11:00	Std
	10 S05	43815	-		10 SFAM01.1	PPB	1/19/2023 11:02	Std
ICV007	ICV007	17041	3.8107	-	SFAM01.1	PPB	1/19/2023 11:05	SMPL
ICB007	ICB007	949	0.0729	-	SFAM01.1	PPB	1/19/2023 11:07	SMPL
CCV076	CCV076	21374	4.8171	-	SFAM01.1	PPB	1/19/2023 11:10	SMPL
CCB076	CCB076	835	0.0465	-	SFAM01.1	PPB	1/19/2023 11:12	SMPL
O1116-11DLX2	YKBM8	28043	6.3661	-	SFAM01.1	PPB	1/19/2023 11:14	SMPL
CCV077	CCV077	21014	4.7335	-	SFAM01.1	PPB	1/19/2023 11:19	SMPL
CCB077	CCB077	813	0.0414	-	SFAM01.1	PPB	1/19/2023 11:21	SMPL