


L3123740

7471B

INSTRUMENT ID: CY1

Linear



A= 0.0000e+000

B= 1.9682e-004

C= -3.2466e-003

Rho= 0.9999816

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/00
0.0	0.000	-0.000	-0.000	14	0.000	14					
0.2	0.200	0.199	-0.001	1026	0.0 %	1026					
2.5	2.500	2.472	-0.028	12577	0.0 %	12577					
5.0	5.000	5.040	0.040	25622	0.0 %	25622					
7.5	7.500	7.512	0.012	38185	0.0 %	38185					
10.0	10.000	9.978	-0.022	50710	0.0 %	50710					

LB123740 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Jan 2023 12:09:15

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SP	14	-	0.0000	7471B	PPB	19 Jan 2023 12:11:50	S	Std
0.2 - 1	SP-2	1026	-	0.2000	7471B	PPB	19 Jan 2023 12:14:07	S	Std
2.5 - 1	SP-5	12577	-	2.5000	7471B	PPB	19 Jan 2023 12:16:24	S	Std
5.0 - 1	SP-5	25622	-	5.0000	7471B	PPB	19 Jan 2023 12:18:41	S	Std
7.5 - 1	SP-5	38185	-	7.5000	7471B	PPB	19 Jan 2023 12:20:58	S	Std
10.0 - 1	SP-5	50710	-	10.0000	7471B	PPB	19 Jan 2023 12:23:14	S	Std
ICV08 - 1	ICV08	20793	4.0893	-	7471B	PPB	19 Jan 2023 12:25:59	U	SMPL
ICB08 - 1	ICB08	222	0.0404	-	7471B	PPB	19 Jan 2023 12:28:13	U	SMPL
CCV78 - 1	CCV78	25731	5.0612	-	7471B	PPB	19 Jan 2023 12:30:29	U	SMPL
CCB78 - 1	CCB78	231	0.0422	-	7471B	PPB	19 Jan 2023 12:32:44	U	SMPL
CRA - 1	CRA	1021	0.1977	-	7471B	PPB	19 Jan 2023 12:34:59	U	SMPL
HighStd - 1	HighStd	50789	9.9931	-	7471B	PPB	19 Jan 2023 12:37:14	U	SMPL
ChkStd - 1	ChkStd	35633	7.0101	-	7471B	PPB	19 Jan 2023 12:39:29	U	SMPL
PB150347BL - 1	PBS	-44	-0.0119	-	7471B	PPB	19 Jan 2023 12:41:45	U	SMPL
PB150347BS - 1	LCSS	18789	3.8948	-	7471B	PPB	19 Jan 2023 12:44:01	U	SMPL
O1175-01 - 1	YARD-1	1607	0.3130	-	7471B	PPB	19 Jan 2023 12:46:17	U	SMPL
O1175-10 - 1	YARD-2	1441	0.2804	-	7471B	PPB	19 Jan 2023 12:48:34	U	SMPL
O1178-01 - 1	AU-5-011723	1533	0.2985	-	7471B	PPB	19 Jan 2023 12:50:51	U	SMPL
O1179-01 - 1	VNJ-208	9750	1.9158	-	7471B	PPB	19 Jan 2023 12:53:07	U	SMPL
O1179-03 - 1	RB21127	8622	1.6937	-	7471B	PPB	19 Jan 2023 12:55:23	U	SMPL
CCV79 - 1	CCV79	25220	4.9606	-	7471B	PPB	19 Jan 2023 12:57:38	U	SMPL
CCB79 - 1	CCB79	195	0.0351	-	7471B	PPB	19 Jan 2023 12:59:53	U	SMPL
O1179-03DUP - 1	RB21127DUP	9468	1.8603	-	7471B	PPB	19 Jan 2023 13:02:09	U	SMPL
O1179-03MS - 1	RB21127MS	25880	5.0905	-	7471B	PPB	19 Jan 2023 13:04:24	U	SMPL
O1179-03MSD - 1	RB21127MSD	27192	5.3487	-	7471B	PPB	19 Jan 2023 13:06:40	U	SMPL
O1180-01 - 1	MH-7	557	0.1064	-	7471B	PPB	19 Jan 2023 13:08:56	U	SMPL
O1198-01 - 1	PL-02-011823	826	0.1593	-	7471B	PPB	19 Jan 2023 13:11:12	U	SMPL
O1214-02 - 1	TJRC-011723-B3-0-10	4596	0.9013	-	7471B	PPB	19 Jan 2023 13:13:28	U	SMPL
O1214-05 - 1	TJRC-011723-B3-10-20	180	0.0322	-	7471B	PPB	19 Jan 2023 13:15:45	U	SMPL
O1214-08 - 1	TJRC-011723-B3-0-10-DUP	4550	0.8923	-	7471B	PPB	19 Jan 2023 13:18:02	U	SMPL
O1214-11 - 1	TJRC-011823-B3-20-30	105	0.0174	-	7471B	PPB	19 Jan 2023 13:20:20	U	SMPL
O1214-14 - 1	TJRC-011823-B3-30-40	78	0.0121	-	7471B	PPB	19 Jan 2023 13:22:38	U	SMPL
CCV80 - 1	CCV80	25831	5.0808	-	7471B	PPB	19 Jan 2023 13:24:53	U	SMPL
CCB80 - 1	CCB80	144	0.0251	-	7471B	PPB	19 Jan 2023 13:27:09	U	SMPL
O1215-01 - 1	BU-02-011823	5061	0.9929	-	7471B	PPB	19 Jan 2023 13:29:25	U	SMPL
O1217-01 - 1	BORING-4	197583	38.8853	-	7471B	PPB	19 Jan 2023 13:31:40	U	SMPL
O1217-14 - 1	BORING-2	1928	0.3762	-	7471B	PPB	19 Jan 2023 13:33:56	U	SMPL
O1220-01 - 1	MH-8	748	0.1440	-	7471B	PPB	19 Jan 2023 13:36:15	U	SMPL
O1220-04 - 1	TP-G	440	0.0834	-	7471B	PPB	19 Jan 2023 13:38:31	U	SMPL
O1179-03LX5 - 1		1880	0.3668	-	7471B	PPB	19 Jan 2023 13:40:48	U	SMPL
O1179-03A - 1		30534	6.0163	-	7471B	PPB	19 Jan 2023 13:43:04	U	SMPL
O1217-01DLX10 - 1		21617	4.2514	-	7471B	PPB	19 Jan 2023 13:45:21	U	SMPL
CCV81 - 1	CCV81	25943	5.1029	-	7471B	PPB	19 Jan 2023 13:47:38	U	SMPL
CCB81 - 1	CCB81	210	0.0381	-	7471B	PPB	19 Jan 2023 13:49:53	U	SMPL