

LB123749


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42065



A= 0.0000e+000

B= 2.3984e-004

C= 1.3906e-002

Rho= 0.9997702

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.014	0.014	-1	0.000	-1					-
0.2	0.200	0.190	-0.010	733	0.0 %	733					-5
2.5	2.500	2.566	0.066	10640	0.0 %	10640					3
5.0	5.000	4.975	-0.025	20686	0.0 %	20686					-1
7.5	7.500	7.353	-0.147	30598	0.0 %	30598					-2
10.0	10.000	10.103	0.103	42065	0.0 %	42065					1

LB123742 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Jan 2023 14:42:19

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-1	-	0.0000	7470A	PPB	19 Jan 2023 14:54:58	S	Std
0.2 - 1	50.2	733	-	0.2000	7470A	PPB	19 Jan 2023 14:57:16	S	Std
2.5 - 1	50.5	10640	-	2.5000	7470A	PPB	19 Jan 2023 14:59:34	S	Std
5.0 - 1	55	20686	-	5.0000	7470A	PPB	19 Jan 2023 15:01:50	S	Std
7.5 - 1	55.5	30598	-	7.5000	7470A	PPB	19 Jan 2023 15:04:07	S	Std
10.0 - 1	510	42065	-	10.0000	7470A	PPB	19 Jan 2023 15:06:22	S	Std
ICV09 - 1	ICV09	15597	3.7547	-	7470A	PPB	19 Jan 2023 15:09:20	U	SMPL
ICB09 - 1	ICB09	18	0.0182	-	7470A	PPB	19 Jan 2023 15:11:34	U	SMPL
CCV82 - 1	CCV82	20581	4.9501	-	7470A	PPB	19 Jan 2023 15:13:49	U	SMPL
CCB82 - 1	CCB82	-5	0.0127	-	7470A	PPB	19 Jan 2023 15:16:05	U	SMPL
CRA - 1	CRA	752	0.1943	-	7470A	PPB	19 Jan 2023 15:18:20	U	SMPL
HighStd - 1	HighStd	39089	9.3891	-	7470A	PPB	19 Jan 2023 15:20:35	U	SMPL
ChkStd - 1	ChkStd	27978	6.7242	-	7470A	PPB	19 Jan 2023 15:22:51	U	SMPL
PB150327BL - 1	PBW	-58	-0.0000	-	7470A	PPB	19 Jan 2023 15:25:06	U	SMPL
PB150327BS - 1	LCSW	14748	3.5511	-	7470A	PPB	19 Jan 2023 15:27:22	U	SMPL
O1161-04 - 1	20230113-H4-STOCKPILE	9	0.0161	-	7470A	PPB	19 Jan 2023 15:29:39	U	SMPL
O1172-02 - 1	TP-F	19	0.0185	-	7470A	PPB	19 Jan 2023 15:31:55	U	SMPL
O1175-02 - 1	YARD-1	21	0.0189	-	7470A	PPB	19 Jan 2023 15:34:12	U	SMPL
O1175-11 - 1	YARD-2	2	0.0144	-	7470A	PPB	19 Jan 2023 15:36:29	U	SMPL
O1180-02 - 1	MH-7	-4	0.0129	-	7470A	PPB	19 Jan 2023 15:38:44	U	SMPL
CCV83 - 1	CCV83	20640	4.9643	-	7470A	PPB	19 Jan 2023 15:40:59	U	SMPL
CCB83 - 1	CCB83	-47	0.0026	-	7470A	PPB	19 Jan 2023 15:43:15	U	SMPL
O1180-02DUP - 1	MH-7DUP	15	0.0175	-	7470A	PPB	19 Jan 2023 15:45:30	U	SMPL
O1180-02MS - 1	MH-7MS	14548	3.5031	-	7470A	PPB	19 Jan 2023 15:47:45	U	SMPL
O1180-02MSD - 1	MH-7MSD	14027	3.3782	-	7470A	PPB	19 Jan 2023 15:50:01	U	SMPL
O1182-02 - 1	002	49	0.0257	-	7470A	PPB	19 Jan 2023 15:52:16	U	SMPL
PB150328BL - 1	PBW	22	0.0192	-	7470A	PPB	19 Jan 2023 15:54:32	U	SMPL
PB150328BS - 1	LCSW	14395	3.4664	-	7470A	PPB	19 Jan 2023 15:56:49	U	SMPL
O1186-01 - 1	MW-UNK	3	0.0146	-	7470A	PPB	19 Jan 2023 15:59:05	U	SMPL
O1186-02 - 1	MW-UNK	23	0.0194	-	7470A	PPB	19 Jan 2023 16:01:22	U	SMPL
O1186-03 - 1	TWP04	245	0.0727	-	7470A	PPB	19 Jan 2023 16:03:39	U	SMPL
O1186-04 - 1	TWP04	27	0.0204	-	7470A	PPB	19 Jan 2023 16:05:54	U	SMPL
CCV84 - 1	CCV84	20525	4.9367	-	7470A	PPB	19 Jan 2023 16:08:09	U	SMPL
CCB84 - 1	CCB84	-46	0.0029	-	7470A	PPB	19 Jan 2023 16:10:25	U	SMPL
O1209-01 - 1	TWP-03	9649	2.3281	-	7470A	PPB	19 Jan 2023 16:12:41	U	SMPL
O1209-01DUP - 1	TWP-03DUP	9218	2.2248	-	7470A	PPB	19 Jan 2023 16:14:56	U	SMPL
O1209-01MS - 1	TWP-03MS	24987	6.0069	-	7470A	PPB	19 Jan 2023 16:17:12	U	SMPL
O1209-01MSD - 1	TWP-03MSD	25004	6.0109	-	7470A	PPB	19 Jan 2023 16:22:12	U	SMPL
O1209-02 - 1	TWP-03	-52	0.0014	-	7470A	PPB	19 Jan 2023 16:24:28	U	SMPL
PB150307TB - 1	PB150307TB	-34	0.0058	-	7470A	PPB	19 Jan 2023 16:26:45	U	SMPL
O1180-02LX5 - 1		-26	0.0077	-	7470A	PPB	19 Jan 2023 16:29:01	U	SMPL
O1180-02A - 1		14985	3.6079	-	7470A	PPB	19 Jan 2023 16:31:18	U	SMPL
O1209-01LX5 - 1		2043	0.5039	-	7470A	PPB	19 Jan 2023 16:33:35	U	SMPL
O1209-01A - 1		26076	6.2680	-	7470A	PPB	19 Jan 2023 16:35:51	U	SMPL
CCV85 - 1	CCV85	19952	4.7992	-	7470A	PPB	19 Jan 2023 16:38:06	U	SMPL
CCB85 - 1	CCB85	-51	0.0017	-	7470A	PPB	19 Jan 2023 16:40:22	U	SMPL