



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: RUBINA

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB123744

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: FLASHPOINT

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2744

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.7	01/20/2023	09:00
2	O1236-01		1	12	100.00	>212.0	>212.0	01/20/2023	09:30
3	O1236-03		1	14	100.00	>212.0	>212.0	01/20/2023	10:00
4	O1236-03DUP		1	13	100.00	>212.0	>212.0	01/20/2023	10:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp1-20

WorkList ID : 166707

Department : Wet-Chemistry

Date : 01-20-2023 08:17:45

6123744

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/26/2023	Water	O1236-01	Flash Point	Cool 4 deg C	PSEG03	N12	NWB-1875	01/19/2023	1010B
01/26/2023	Water	O1236-03	Flash Point	Cool 4 deg C	PSEG03	N12	NWB-1770	01/19/2023	1010B

Date/Time 01/26/2023 08:45
 Raw Sample Received by: RM CWS
 Raw Sample Relinquished by: JDC

Date/Time 01/26/2023 11:20
 Raw Sample Received by: JDC
 Raw Sample Relinquished by: RM CWS