

LB 123749


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42133



A= 0.0000e+000

B= 2.3853e-004

C= -2.2941e-002

Rho= 0.9999144

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/μ
0.0	0.000	-0.015	-0.015	33	0.000	33					
0.2	0.200	0.172	-0.028	819	0.0 %	819					-14
2.5	2.500	2.558	0.058	10822	0.0 %	10822					2
5.0	5.000	5.042	0.042	21232	0.0 %	21232					1
7.5	7.500	7.416	-0.084	31185	0.0 %	31185					-1
10.0	10.000	10.027	0.027	42133	0.0 %	42133					0

LB123749 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 20 Jan 2023 12:12:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	33	-	0.0000	7470A	PPB	20 Jan 2023 12:23:26	S	Std
0.2 - 1	50.2	819	-	0.2000	7470A	PPB	20 Jan 2023 12:25:43	S	Std
2.5 - 1	52.5	10822	-	2.5000	7470A	PPB	20 Jan 2023 12:28:00	S	Std
5.0 - 1	53	21232	-	5.0000	7470A	PPB	20 Jan 2023 12:30:18	S	Std
7.5 - 1	53.5	31185	-	7.5000	7470A	PPB	20 Jan 2023 12:32:34	S	Std
10.0 - 1	510	42133	-	10.0000	7470A	PPB	20 Jan 2023 12:37:24	S	Std
ICV10 - 1	ICV10	16544	3.9233	-	7470A	PPB	20 Jan 2023 12:42:37	U	SMPL
ICB10 - 1	ICB10	39	-0.0136	-	7470A	PPB	20 Jan 2023 12:44:52	U	SMPL
CCV86 - 1	CCV86	20370	4.8359	-	7470A	PPB	20 Jan 2023 12:47:07	U	SMPL
CCB86 - 1	CCB86	13	-0.0198	-	7470A	PPB	20 Jan 2023 12:49:22	U	SMPL
CRA - 1	CRA	927	0.1982	-	7470A	PPB	20 Jan 2023 12:54:11	U	SMPL
HighStd - 1	HighStd	41610	9.9023	-	7470A	PPB	20 Jan 2023 12:56:26	U	SMPL
ChkStd - 1	ChkStd	28144	6.6903	-	7470A	PPB	20 Jan 2023 12:58:42	U	SMPL
PB150351BL - 1	PBW	-27	-0.0294	-	7470A	PPB	20 Jan 2023 13:00:58	U	SMPL
PB150351BS - 1	LCSW	15785	3.7423	-	7470A	PPB	20 Jan 2023 13:03:14	U	SMPL
O1161-04 - 1	20230113-H4-STOCKPILE	65	-0.0074	-	7470A	PPB	20 Jan 2023 13:05:31	U	SMPL
O1200-01 - 1	C-1	8	-0.0210	-	7470A	PPB	20 Jan 2023 13:07:48	U	SMPL
O1214-03 - 1	TJRC-011723-B3-0-10	2	-0.0225	-	7470A	PPB	20 Jan 2023 13:10:05	U	SMPL
O1214-06 - 1	TJRC-011723-B3-10-20	-15	-0.0265	-	7470A	PPB	20 Jan 2023 13:12:21	U	SMPL
O1214-09 - 1	TJRC-011723-B3-0-10-DUP	4	-0.0220	-	7470A	PPB	20 Jan 2023 13:14:37	U	SMPL
CCV87 - 1	CCV87	20360	4.8335	-	7470A	PPB	20 Jan 2023 13:16:52	U	SMPL
CCB87 - 1	CCB87	-43	-0.0332	-	7470A	PPB	20 Jan 2023 13:19:07	U	SMPL
O1214-12 - 1	TJRC-011823-B3-20-30	15	-0.0194	-	7470A	PPB	20 Jan 2023 13:21:22	U	SMPL
O1214-15 - 1	TJRC-011823-B3-30-40	38	-0.0139	-	7470A	PPB	20 Jan 2023 13:23:38	U	SMPL
O1217-02 - 1	BORING-4	85	-0.0027	-	7470A	PPB	20 Jan 2023 13:25:54	U	SMPL
O1217-15 - 1	BORING-2	28	-0.0163	-	7470A	PPB	20 Jan 2023 13:28:09	U	SMPL
O1220-02 - 1	MH-8	35	-0.0146	-	7470A	PPB	20 Jan 2023 13:30:26	U	SMPL
O1220-05 - 1	TP-G	38	-0.0139	-	7470A	PPB	20 Jan 2023 13:32:42	U	SMPL
O1220-05DUP - 1	TP-GDUP	58	-0.0091	-	7470A	PPB	20 Jan 2023 13:34:58	U	SMPL
O1220-05MS - 1	TP-GMS	15929	3.7766	-	7470A	PPB	20 Jan 2023 13:37:15	U	SMPL
O1220-05MSD - 1	TP-GMSD	18004	4.2716	-	7470A	PPB	20 Jan 2023 13:39:32	U	SMPL
PB150353BL - 1	PBW	34	-0.0148	-	7470A	PPB	20 Jan 2023 13:41:51	U	SMPL
CCV88 - 1	CCV88	20320	4.8240	-	7470A	PPB	20 Jan 2023 13:44:08	U	SMPL
CCB88 - 1	CCB88	61	-0.0084	-	7470A	PPB	20 Jan 2023 13:46:24	U	SMPL
PB150353BS - 1	LCSW	15145	3.5896	-	7470A	PPB	20 Jan 2023 13:48:39	U	SMPL
O1199-01 - 1	34185	361	0.0632	-	7470A	PPB	20 Jan 2023 13:50:55	U	SMPL
O1214-18 - 1	TJRC-011823-GW	126	0.0071	-	7470A	PPB	20 Jan 2023 13:53:11	U	SMPL
O1214-19 - 1	TJRC-011823-GW-DUP	124	0.0066	-	7470A	PPB	20 Jan 2023 13:55:26	U	SMPL
O1221-01 - 1	MW-10-0118	119	0.0054	-	7470A	PPB	20 Jan 2023 13:57:44	U	SMPL
O1221-01DUP - 1	MW-10-0118DUP	120	0.0057	-	7470A	PPB	20 Jan 2023 14:00:00	U	SMPL
O1221-01MS - 1	MW-10-0118MS	14087	3.3372	-	7470A	PPB	20 Jan 2023 14:02:16	U	SMPL
O1221-01MSD - 1	MW-10-0118MSD	15122	3.5841	-	7470A	PPB	20 Jan 2023 14:04:33	U	SMPL
O1221-02 - 1	MW-11-0118	10	-0.0206	-	7470A	PPB	20 Jan 2023 14:06:51	U	SMPL
O1221-03 - 1	MW-12-0118	63	-0.0079	-	7470A	PPB	20 Jan 2023 14:09:06	U	SMPL
CCV89 - 1	CCV89	21824	5.1827	-	7470A	PPB	20 Jan 2023 14:11:22	U	SMPL
CCB89 - 1	CCB89	-26	-0.0291	-	7470A	PPB	20 Jan 2023 14:13:38	U	SMPL
O1221-04 - 1	MW-13-0118	28	-0.0163	-	7470A	PPB	20 Jan 2023 14:15:53	U	SMPL
O1236-01 - 1	NWB-1875	54	-0.0101	-	7470A	PPB	20 Jan 2023 14:18:09	U	SMPL
PB150329TB - 1	PB150329TB	-51	-0.0351	-	7470A	PPB	20 Jan 2023 14:20:25	U	SMPL
O1220-05LX5 - 1		-42	-0.0330	-	7470A	PPB	20 Jan 2023 14:22:40	U	SMPL
O1220-05A - 1		19441	4.6143	-	7470A	PPB	20 Jan 2023 14:24:56	U	SMPL
O1221-01LX5 - 1		-90	-0.0444	-	7470A	PPB	20 Jan 2023 14:27:13	U	SMPL
O1221-01A - 1		17107	4.0576	-	7470A	PPB	20 Jan 2023 14:29:29	U	SMPL
CCV90 - 1	CCV90	21741	5.1629	-	7470A	PPB	20 Jan 2023 14:31:47	U	SMPL
CCB90 - 1	CCB90	-45	-0.0337	-	7470A	PPB	20 Jan 2023 14:34:04	U	SMPL