



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB123750

Analyst By : jignesh
Supervisor Review By : Iwona
Slope : 98.6
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	01/20/2023	12:20
2	CAL2	1	Water	NA	NA	20.7	7.00	01/20/2023	12:21
3	CAL3	1	Water	NA	NA	20.8	10.03	01/20/2023	12:22
4	ICV	1	Water	NA	NA	20.8	6.99	01/20/2023	12:25
5	CCV1	1	Water	NA	NA	20.8	2.01	01/20/2023	12:26
6	O1235-01	1	Water	NA	NA	20.6	6.43	01/20/2023	12:30
7	O1235-01DUP	1	Water	NA	NA	20.7	6.44	01/20/2023	12:31
8	O1235-02	1	Water	NA	NA	20.7	6.22	01/20/2023	12:33
9	O1235-03	1	Water	NA	NA	20.6	6.12	01/20/2023	12:35
10	O1235-08	1	Water	NA	NA	20.5	6.97	01/20/2023	12:36
11	O1235-09	1	Water	NA	NA	20.2	7.60	01/20/2023	12:37
12	O1235-10	1	Water	NA	NA	20.2	7.10	01/20/2023	12:40
13	O1235-13	1	Water	NA	NA	20.2	7.10	01/20/2023	12:45
14	O1235-14	1	Water	NA	NA	20.2	7.10	01/20/2023	12:47
15	O1236-01	1	Water	NA	NA	20.9	4.49	01/20/2023	12:50
16	CCV2	1	Water	NA	NA	20.8	12.03	01/20/2023	12:54
17	O1236-03	1	Water	NA	NA	22.5	7.60	01/20/2023	13:00
18	CCV3	1	Water	NA	NA	20.8	2.01	01/20/2023	13:05

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 01235 w

WorkList ID : 166735

Department : Wet-Chemistry

Date : 01-20-2023 12:09:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/18/2023	Water	O1235-01	pH	A	Cool 4 deg C	ATCE02	N11	12X49	01/18/2023 9040C
01/18/2023	Water	O1235-02	pH	A	Cool 4 deg C	ATCE02	N11	12X50	01/18/2023 9040C
01/18/2023	Water	O1235-03	pH	A	Cool 4 deg C	ATCE02	N11	12X51	01/18/2023 9040C
01/18/2023	Water	O1235-08	pH	A	Cool 4 deg C	ATCE02	N11	12X56	01/18/2023 9040C
01/18/2023	Water	O1235-09	pH	A	Cool 4 deg C	ATCE02	N11	12X57	01/18/2023 9040C
01/18/2023	Water	O1235-10	pH	A	Cool 4 deg C	ATCE02	N11	12X58	01/18/2023 9040C
01/18/2023	Water	O1235-13	pH	A	Cool 4 deg C	ATCE02	N11	12X59	01/18/2023 9040C
01/18/2023	Water	O1235-14	pH	A	Cool 4 deg C	ATCE02	N11	12X48	01/18/2023 9040C
01/19/2023	Water	O1236-01	pH	A	Cool 4 deg C	PSEG03	N12	NWB-1875	01/19/2023 9040C
01/19/2023	Water	O1236-03	pH	C	Cool 4 deg C	PSEG03	N12	NWB-1770	01/19/2023 9040C

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by: