

LB123755


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

41700



A= 0.0000e+000

B= 2.4018e-004 slope

C= 2.6261e-002 y-intercept

Rho= 0.9999499

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.032	0.032	25	0.000	25					
0.05	0.050					185	0				
0.20	0.200	0.209	0.009	760	0.0 %	760					
2.50	2.500	2.453	-0.047	10102	0.0 %	10102					
5.00	5.000	5.017	0.017	20781	0.0 %	20781					
7.50	7.500	7.447	-0.053	30898	0.0 %	30898					
10.0	10.000	10.042	0.042	41700	0.0 %	41700					

LB123755

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0	25	-		0 SFAM01.1	PPB	1/20/2023 16:13	Std
	0.2 S01	760	-		0.2 SFAM01.1	PPB	1/20/2023 16:20	Std
	2.5 S02	10102	-		2.5 SFAM01.1	PPB	1/20/2023 16:22	Std
	5 S03	20781	-		5 SFAM01.1	PPB	1/20/2023 16:25	Std
	7.5 S04	30898	-		7.5 SFAM01.1	PPB	1/20/2023 16:27	Std
	10 S05	41700	-		10 SFAM01.1	PPB	1/20/2023 16:29	Std
ICV011	ICV011	16328	3.9479 -		SFAM01.1	PPB	1/20/2023 16:32	SMPL
ICB011	ICB011	20	0.0311 -		SFAM01.1	PPB	1/20/2023 16:34	SMPL
CCV091	CCV091	20701	4.9982 -		SFAM01.1	PPB	1/20/2023 16:37	SMPL
CCB091	CCB091	-3	0.0255 -		SFAM01.1	PPB	1/20/2023 16:39	SMPL
PB150377BL	PBW377	9	0.0284 -		SFAM01.1	PPB	1/20/2023 16:41	SMPL
O1128-07	X3E32	26720	6.4438 -		SFAM01.1	PPB	1/20/2023 16:46	SMPL
O1128-08	X3E32D	27257	6.5728 -		SFAM01.1	PPB	1/20/2023 16:48	SMPL
O1128-09	X3E32S	31340	7.5534 -		SFAM01.1	PPB	1/20/2023 16:50	SMPL
CCV092	CCV092	20416	4.9297 -		SFAM01.1	PPB	1/20/2023 16:52	SMPL
CCB092	CCB092	-1	0.026 -		SFAM01.1	PPB	1/20/2023 16:55	SMPL