



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 1/24/2023

OVENTEMP IN Celsius(°C): 107
Time IN: 16:40
In Date: 01/23/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:10
Out Date: 01/24/2023
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB123765

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
O1251-15	EPH-3	1	1.14	8.69	9.83	8.54	85.2	
O1260-01	TP-21	2	1.13	8.45	9.58	9.44	98.3	
O1260-03	TP-21-EPH-2	3	1.15	8.67	9.82	9.32	94.2	
O1260-04	TP-22	4	1.14	8.69	9.83	9.2	92.8	
O1260-06	TP-22-EPH-2	5	1.14	8.58	9.72	9.17	93.6	
O1260-07	TP-23	6	1.15	8.83	9.98	9.55	95.1	
O1260-09	TP-23-EPH-2	7	1.13	8.73	9.86	9.46	95.4	
O1260-10	TP-24	8	1.14	8.83	9.97	9.63	96.1	
O1260-12	TP-24-EPH-2	9	1.14	8.84	9.98	9.66	96.4	
O1260-13	TP-25	10	1.14	8.78	9.92	9.48	95.0	
O1260-15	TP-25-EPH-2	11	1.13	8.85	9.98	9.7	96.8	
O1260-16	TP-26	12	1.13	8.56	9.69	9.22	94.5	
O1260-18	TP-26-EPH-2	13	1.13	8.79	9.92	9.46	94.8	
O1260-19	TP-27	14	1.13	8.73	9.86	9.4	94.7	
O1260-21	TP-27-EPH-2	15	1.12	8.68	9.8	9.36	94.9	
O1260-22	TP-28	16	1.13	8.41	9.54	9.07	94.4	
O1260-24	TP-28-EPH-2	17	1.14	8.43	9.57	9.12	94.7	
O1266-01	VNJ-219	19	1.14	8.49	9.63	8.2	83.2	
O1267-01	01A-01B-01C	18	1.00	1.00	2.00	2.00	100.0	CAULKING SAMPLE
O1268-01	TR-06-012323	20	1.11	8.61	9.72	8.66	87.7	
O1268-02	TR-06-012323-EPH-2	21	1.11	8.61	9.72	8.66	87.7	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$