

7470A


LB123775

INSTRUMENT ID: CV2

Linear

 $\mu$  Abs.:

37635



A= 0.0000e+000

B= 2.6569e-004

C= 1.1532e-002

Rho= 0.9993269

Accept=Accepted

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|----|
| 0.0    | 0.000  | 0.008  | 0.008  | -13   | 0.000      | -13   |       |       |       |       |    |
| 0.2    | 0.200  | 0.185  | -0.015 | 652   | 0.0 %      | 652   |       |       |       |       | -8 |
| 2.5    | 2.500  | 2.405  | -0.095 | 9008  | 0.0 %      | 9008  |       |       |       |       | -4 |
| 5.0    | 5.000  | 5.265  | 0.265  | 19774 | 0.0 %      | 19774 |       |       |       |       | 5  |
| 7.5    | 7.500  | 7.326  | -0.174 | 27532 | 0.0 %      | 27532 |       |       |       |       | 2  |
| 10.0   | 10.000 | 10.011 | 0.011  | 37635 | 0.0 %      | 37635 |       |       |       |       | 0  |

## LB123775 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Jan 2023 10:15:05

| Sample ID       | Extended ID           | $\mu$ Abs. | Conc.   | Std Conc | Method | Units | Date                 | Type | Type |
|-----------------|-----------------------|------------|---------|----------|--------|-------|----------------------|------|------|
| 0.0 - 1         | 50                    | -13        | -       | 0.0000   | 7470A  | PPB   | 24 Jan 2023 10:25:37 | S    | Std  |
| 0.2 - 1         | 50.2                  | 652        | -       | 0.2000   | 7470A  | PPB   | 24 Jan 2023 10:27:55 | S    | Std  |
| 2.5 - 1         | 52.5                  | 9008       | -       | 2.5000   | 7470A  | PPB   | 24 Jan 2023 10:30:12 | S    | Std  |
| 5.0 - 1         | 55                    | 19774      | -       | 5.0000   | 7470A  | PPB   | 24 Jan 2023 10:35:02 | S    | Std  |
| 7.5 - 1         | 57.5                  | 27532      | -       | 7.5000   | 7470A  | PPB   | 24 Jan 2023 10:37:18 | S    | Std  |
| 10.0 - 1        | 510                   | 37635      | -       | 10.0000  | 7470A  | PPB   | 24 Jan 2023 10:42:09 | S    | Std  |
| ICV15 - 1       | ICV15                 | 15040      | 4.0075  | -        | 7470A  | PPB   | 24 Jan 2023 10:47:45 | U    | SMPL |
| ICB15 - 1       | ICB15                 | -6         | 0.0099  | -        | 7470A  | PPB   | 24 Jan 2023 10:49:59 | U    | SMPL |
| CCV08 - 1       | CCV08                 | 18707      | 4.9817  | -        | 7470A  | PPB   | 24 Jan 2023 10:52:15 | U    | SMPL |
| CCB08 - 1       | CCB08                 | -57        | -0.0036 | -        | 7470A  | PPB   | 24 Jan 2023 10:54:30 | U    | SMPL |
| CRA - 1         | CRA                   | 629        | 0.1786  | -        | 7470A  | PPB   | 24 Jan 2023 10:59:19 | U    | SMPL |
| HighStd - 1     | HighStd               | 35578      | 9.4641  | -        | 7470A  | PPB   | 24 Jan 2023 11:01:34 | U    | SMPL |
| ChkStd - 1      | ChkStd                | 23526      | 6.2621  | -        | 7470A  | PPB   | 24 Jan 2023 11:03:50 | U    | SMPL |
| PB150390BL - 1  | PBW                   | -138       | -0.0251 | -        | 7470A  | PPB   | 24 Jan 2023 11:06:06 | U    | SMPL |
| PB150390BS - 1  | LCSW                  | 14544      | 3.8757  | -        | 7470A  | PPB   | 24 Jan 2023 11:10:59 | U    | SMPL |
| O1232-03 - 1    | WC-17                 | -75        | -0.0084 | -        | 7470A  | PPB   | 24 Jan 2023 11:13:15 | U    | SMPL |
| O1233-10 - 1    | WC-38                 | -46        | -0.0007 | -        | 7470A  | PPB   | 24 Jan 2023 11:15:32 | U    | SMPL |
| O1235-04 - 1    | 12X52                 | -37        | 0.0017  | -        | 7470A  | PPB   | 24 Jan 2023 11:17:49 | U    | SMPL |
| O1235-05 - 1    | 12X53                 | 23         | 0.0176  | -        | 7470A  | PPB   | 24 Jan 2023 11:20:06 | U    | SMPL |
| O1235-06 - 1    | 12X54                 | 7          | 0.0134  | -        | 7470A  | PPB   | 24 Jan 2023 11:22:22 | U    | SMPL |
| CCV09 - 1       | CCV09                 | 18854      | 5.0208  | -        | 7470A  | PPB   | 24 Jan 2023 11:24:37 | U    | SMPL |
| CCB09 - 1       | CCB09                 | -73        | -0.0079 | -        | 7470A  | PPB   | 24 Jan 2023 11:26:53 | U    | SMPL |
| O1235-07 - 1    | 12X55                 | -4         | 0.0105  | -        | 7470A  | PPB   | 24 Jan 2023 11:29:08 | U    | SMPL |
| O1237-04 - 1    | COMP-1-2-3A           | -1         | 0.0113  | -        | 7470A  | PPB   | 24 Jan 2023 11:31:23 | U    | SMPL |
| O1238-02 - 1    | RBRB22281-OILY-DEBRIS | -1         | 0.0113  | -        | 7470A  | PPB   | 24 Jan 2023 11:33:38 | U    | SMPL |
| O1239-02 - 1    | BORING-3              | -2         | 0.0110  | -        | 7470A  | PPB   | 24 Jan 2023 11:35:54 | U    | SMPL |
| O1251-02 - 1    | COMP-1                | 13         | 0.0150  | -        | 7470A  | PPB   | 24 Jan 2023 11:38:11 | U    | SMPL |
| O1251-08 - 1    | COMP-2                | 4          | 0.0126  | -        | 7470A  | PPB   | 24 Jan 2023 11:40:27 | U    | SMPL |
| O1252-03 - 1    | MH-9                  | 3          | 0.0123  | -        | 7470A  | PPB   | 24 Jan 2023 11:42:44 | U    | SMPL |
| O1255-02 - 1    | BORING-1              | 4          | 0.0126  | -        | 7470A  | PPB   | 24 Jan 2023 11:45:01 | U    | SMPL |
| O1255-02DUP - 1 | BORING-1DUP           | 4          | 0.0126  | -        | 7470A  | PPB   | 24 Jan 2023 11:47:18 | U    | SMPL |
| O1255-02MS - 1  | BORING-1MS            | 16089      | 4.2862  | -        | 7470A  | PPB   | 24 Jan 2023 11:49:34 | U    | SMPL |
| CCV10 - 1       | CCV10                 | 18133      | 4.8292  | -        | 7470A  | PPB   | 24 Jan 2023 11:51:49 | U    | SMPL |
| CCB10 - 1       | CCB10                 | -50        | -0.0018 | -        | 7470A  | PPB   | 24 Jan 2023 11:54:05 | U    | SMPL |
| O1255-02MSD - 1 | BORING-1MSD           | 16108      | 4.2912  | -        | 7470A  | PPB   | 24 Jan 2023 11:56:20 | U    | SMPL |
| PB150391BL - 1  | PBW                   | -19        | 0.0065  | -        | 7470A  | PPB   | 24 Jan 2023 11:58:36 | U    | SMPL |
| PB150391BS - 1  | LCSW                  | 14840      | 3.9543  | -        | 7470A  | PPB   | 24 Jan 2023 12:00:51 | U    | SMPL |
| O1235-01 - 1    | 12X49                 | -21        | 0.0060  | -        | 7470A  | PPB   | 24 Jan 2023 12:03:07 | U    | SMPL |
| O1235-02 - 1    | 12X50                 | 10         | 0.0142  | -        | 7470A  | PPB   | 24 Jan 2023 12:05:23 | U    | SMPL |
| O1235-03 - 1    | 12X51                 | 14         | 0.0153  | -        | 7470A  | PPB   | 24 Jan 2023 12:07:39 | U    | SMPL |
| O1235-08 - 1    | 12X56                 | 34         | 0.0206  | -        | 7470A  | PPB   | 24 Jan 2023 12:09:56 | U    | SMPL |
| O1235-09 - 1    | 12X57                 | 38         | 0.0216  | -        | 7470A  | PPB   | 24 Jan 2023 12:12:13 | U    | SMPL |
| O1235-10 - 1    | 12X58                 | 27         | 0.0187  | -        | 7470A  | PPB   | 24 Jan 2023 12:14:30 | U    | SMPL |
| O1235-13 - 1    | 12X59                 | -21        | 0.0060  | -        | 7470A  | PPB   | 24 Jan 2023 12:16:46 | U    | SMPL |
| CCV11 - 1       | CCV11                 | 18787      | 5.0030  | -        | 7470A  | PPB   | 24 Jan 2023 12:19:02 | U    | SMPL |
| CCB11 - 1       | CCB11                 | -49        | -0.0015 | -        | 7470A  | PPB   | 24 Jan 2023 12:21:17 | U    | SMPL |
| O1235-14 - 1    | 12X48                 | -2         | 0.0110  | -        | 7470A  | PPB   | 24 Jan 2023 12:23:33 | U    | SMPL |
| O1235-14DUP - 1 | 12X48DUP              | 13         | 0.0150  | -        | 7470A  | PPB   | 24 Jan 2023 12:25:49 | U    | SMPL |
| O1235-14MS - 1  | 12X48MS               | 16662      | 4.4384  | -        | 7470A  | PPB   | 24 Jan 2023 12:28:04 | U    | SMPL |
| O1235-14MSD - 1 | 12X48MSD              | 16741      | 4.4594  | -        | 7470A  | PPB   | 24 Jan 2023 12:30:20 | U    | SMPL |
| PB150368TB - 1  | PB150368TB            | -40        | 0.0009  | -        | 7470A  | PPB   | 24 Jan 2023 12:32:36 | U    | SMPL |
| PB150371TB - 1  | PB150371TB            | 7          | 0.0134  | -        | 7470A  | PPB   | 24 Jan 2023 12:34:52 | U    | SMPL |
| O1255-02LX5 - 1 |                       | 7          | 0.0134  | -        | 7470A  | PPB   | 24 Jan 2023 12:37:09 | U    | SMPL |
| O1255-02A - 1   |                       | 16118      | 4.2939  | -        | 7470A  | PPB   | 24 Jan 2023 12:39:26 | U    | SMPL |
| O1235-14LX5 - 1 |                       | -59        | -0.0041 | -        | 7470A  | PPB   | 24 Jan 2023 12:41:42 | U    | SMPL |
| O1235-14A - 1   |                       | 16333      | 4.3510  | -        | 7470A  | PPB   | 24 Jan 2023 12:43:58 | U    | SMPL |
| CCV12 - 1       | CCV12                 | 18933      | 5.0418  | -        | 7470A  | PPB   | 24 Jan 2023 12:46:14 | U    | SMPL |
| CCB12 - 1       | CCB12                 | -86        | -0.0113 | -        | 7470A  | PPB   | 24 Jan 2023 12:48:29 | U    | SMPL |