

LB123782


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

50242



A= 0.0000e+000

B= 1.9838e-004

C= 1.8489e-002

Rho= 0.9999466

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.020	0.020	10	0.000	10					
0.2	0.200	0.221	0.021	1021	0.0 %	1021					11
2.5	2.500	2.454	-0.046	12279	0.0 %	12279					-2
5.0	5.000	4.957	-0.043	24896	0.0 %	24896					-1
7.5	7.500	7.561	0.061	38022	0.0 %	38022					1
10.0	10.000	9.985	-0.015	50242	0.0 %	50242					0

LB123782 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Jan 2023 15:04:52

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	10	-	0.0000	7471B	PPB	24 Jan 2023 15:08:51	S	Std
0.2 - 1	50.2	1021	-	0.2000	7471B	PPB	24 Jan 2023 15:11:08	S	Std
2.5 - 1	52.5	12279	-	2.5000	7471B	PPB	24 Jan 2023 15:13:25	S	Std
5.0 - 1	55	24896	-	5.0000	7471B	PPB	24 Jan 2023 15:15:42	S	Std
7.5 - 1	57.5	38022	-	7.5000	7471B	PPB	24 Jan 2023 15:20:44	S	Std
10.0 - 1	511	50242	-	10.0000	7471B	PPB	24 Jan 2023 15:22:59	S	Std
ICV16 - 1	ICV16	18807	3.7494	-	7471B	PPB	24 Jan 2023 15:25:43	U	SMPL
ICB16 - 1	ICB16	-28	0.0129	-	7471B	PPB	24 Jan 2023 15:27:58	U	SMPL
CCV13 - 1	CCV13	24724	4.9232	-	7471B	PPB	24 Jan 2023 15:30:13	U	SMPL
CCB13 - 1	CCB13	-11	0.0163	-	7471B	PPB	24 Jan 2023 15:32:28	U	SMPL
CRA - 1	CRA	1016	0.2200	-	7471B	PPB	24 Jan 2023 15:34:44	U	SMPL
HighStd - 1	HighStd	50576	10.0517	-	7471B	PPB	24 Jan 2023 15:36:59	U	SMPL
ChkStd - 1	ChkStd	35197	7.0009	-	7471B	PPB	24 Jan 2023 15:39:14	U	SMPL
PB150435BL - 1	PBS	-104	-0.0021	-	7471B	PPB	24 Jan 2023 15:41:31	U	SMPL
O1243-02 - 1	TJRC-11823-B4-0-10	4086	0.8291	-	7471B	PPB	24 Jan 2023 15:46:02	U	SMPL
O1243-02DUP - 1	TJRC-11823-B4-0-10DUP	4044	0.8207	-	7471B	PPB	24 Jan 2023 15:48:19	U	SMPL
O1243-02MS - 1	TJRC-11823-B4-0-10MS	23837	4.7473	-	7471B	PPB	24 Jan 2023 15:54:10	U	SMPL
O1243-02MSD - 1	TJRC-11823-B4-0-10MSD	23861	4.7520	-	7471B	PPB	24 Jan 2023 15:56:27	U	SMPL
O1243-04 - 1	TJRC-11823-B4-10-20	81	0.0348	-	7471B	PPB	24 Jan 2023 15:58:42	U	SMPL
PB150435BS - 1	LCSS	17804	3.5504	-	7471B	PPB	24 Jan 2023 16:01:14	U	SMPL
CCV14 - 1	CCV14	23441	4.6687	-	7471B	PPB	24 Jan 2023 16:03:53	U	SMPL
CCB14 - 1	CCB14	-40	0.0106	-	7471B	PPB	24 Jan 2023 16:06:08	U	SMPL
O1243-06 - 1	TJRC-011923-B4-20-30	174	0.0530	-	7471B	PPB	24 Jan 2023 16:08:24	U	SMPL
O1243-08 - 1	TJRC-011923-B4-30-40	461	0.1099	-	7471B	PPB	24 Jan 2023 16:10:39	U	SMPL
O1251-01 - 1	COMP-1	4371	0.8856	-	7471B	PPB	24 Jan 2023 16:12:54	U	SMPL
O1251-07 - 1	COMP-2	729	0.1631	-	7471B	PPB	24 Jan 2023 16:15:10	U	SMPL
O1253-01 - 1	SU-04-012023	2229	0.4607	-	7471B	PPB	24 Jan 2023 16:17:25	U	SMPL
O1254-01 - 1	HR-04-012023	1934	0.4022	-	7471B	PPB	24 Jan 2023 16:19:42	U	SMPL
O1255-01 - 1	BORING-1	1785	0.3726	-	7471B	PPB	24 Jan 2023 16:21:59	U	SMPL
O1256-01 - 1	OR-02-012023	3786	0.7696	-	7471B	PPB	24 Jan 2023 16:24:15	U	SMPL
O1260-01 - 1	TP-21	380789	HIGH	-	7471B	PPB	24 Jan 2023 16:26:34	X	
O1260-04 - 1	TP-22	53867	10.7046	-	7471B	PPB	24 Jan 2023 16:28:27	U	SMPL
CCV15 - 1	CCV15	23643	4.7088	-	7471B	PPB	24 Jan 2023 16:30:45	U	SMPL
CCB15 - 1	CCB15	43	0.0270	-	7471B	PPB	24 Jan 2023 16:33:01	U	SMPL
O1260-07 - 1	TP-23	2188	0.4525	-	7471B	PPB	24 Jan 2023 16:35:17	U	SMPL
O1260-10 - 1	TP-24	1241	0.2647	-	7471B	PPB	24 Jan 2023 16:37:32	U	SMPL
O1260-13 - 1	TP-25	1046	0.2260	-	7471B	PPB	24 Jan 2023 16:39:47	U	SMPL
O1260-16 - 1	TP-26	1002	0.2173	-	7471B	PPB	24 Jan 2023 16:42:04	U	SMPL
O1260-19 - 1	TP-27	854	0.1879	-	7471B	PPB	24 Jan 2023 16:44:20	U	SMPL
O1260-22 - 1	TP-28	567	0.1310	-	7471B	PPB	24 Jan 2023 16:46:36	U	SMPL
O1266-01 - 1	VNJ-219	1912	0.3978	-	7471B	PPB	24 Jan 2023 16:48:52	U	SMPL
O1268-01 - 1	TR-06-012323	1216	0.2596	-	7471B	PPB	24 Jan 2023 16:51:09	U	SMPL
O1243-02LX5 - 1		748	0.1669	-	7471B	PPB	24 Jan 2023 16:53:26	U	SMPL
O1243-02A - 1		24946	4.9677	-	7471B	PPB	24 Jan 2023 16:55:42	U	SMPL
CCV16 - 1	CCV16	24012	4.7820	-	7471B	PPB	24 Jan 2023 17:00:36	U	SMPL
CCB16 - 1	CCB16	-48	0.0090	-	7471B	PPB	24 Jan 2023 17:02:51	U	SMPL
O1260-01DLX10 - 1		53384	10.6088	-	7471B	PPB	24 Jan 2023 17:05:07	U	SMPL
O1260-04DLX2 - 1		27452	5.4644	-	7471B	PPB	24 Jan 2023 17:07:23	U	SMPL
O1260-01DLX25 - 1		18651	3.7188	-	7471B	PPB	24 Jan 2023 17:09:39	U	SMPL
CCV17 - 1	CCV17	23266	4.6340	-	7471B	PPB	24 Jan 2023 17:11:56	U	SMPL
CCB17 - 1	CCB17	-36	0.0116	-	7471B	PPB	24 Jan 2023 17:14:12	U	SMPL