



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB123786

Slope : 99.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.8	4.01	01/24/2023	17:10
2	CAL2	1	Water	NA	NA	20.8	7.00	01/24/2023	17:11
3	CAL3	1	Water	NA	NA	20.7	10.03	01/24/2023	17:12
4	ICV	1	Water	NA	NA	20.7	6.98	01/24/2023	17:15
5	CCV1	1	Water	NA	NA	20.7	2.01	01/24/2023	17:20
6	O1270-02	1	Solid	20.02	20	20.7	10.96	01/24/2023	17:25
7	O1283-02	1	Solid	20.01	20	20.5	10.19	01/24/2023	17:32
8	O1283-05	1	Solid	20.03	20	20.1	7.80	01/24/2023	17:36
9	O1283-05DUP	1	Solid	20.04	20	20.2	7.81	01/24/2023	17:37
10	CCV2	1	Water	NA	NA	20.8	12.03	01/24/2023	17:40

123786

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIIVITY O1283 WorkList ID : 166807 Department : Wet-Chemistry Date : 01-24-2023 15:22:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/23/2023	Solid	O1270-02	Corrosivity	Cool 4 deg C	RUTW01	O11	JC-1-012323	01/23/2023	9045D
01/24/2023	Solid	O1283-02	Corrosivity	Cool 4 deg C	PSEG03	O11	TP-H	01/24/2023	9045D
01/24/2023	Solid	O1283-05	Corrosivity	Cool 4 deg C	PSEG03	O11	MH-10	01/24/2023	9045D

Date/Time 01-24-23 16:40
Raw Sample Received by: JPLWC
Raw Sample Relinquished by: CP sm

Date/Time 01-24-23 18:00
Raw Sample Received by: CP sm
Raw Sample Relinquished by: JPLWC