

7470A


LB123791

INSTRUMENT ID: CV1

Linear

 μ Abs.:

39101



A= 0.0000e+000

B= 2.5812e-004

C= -1.7314e-002

Rho= 0.9996731

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0 D
0.0	0.000	-0.019	-0.019	-8	0.000	-8					
0.2	0.200	0.158	-0.042	680	0.0 %	680					-21
2.5	2.500	2.664	0.164	10388	0.0 %	10388					7
5.0	5.000	4.949	-0.051	19240	0.0 %	19240					-1
7.5	7.500	7.373	-0.127	28632	0.0 %	28632					-2
10.0	10.000	10.075	0.075	39101	0.0 %	39101					1

LB123791 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 Jan 2023 10:13:43

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-8	-	0.0000	7470A	PPB	25 Jan 2023 10:29:42	S	Std
0.2 - 1	50.2	680	-	0.2000	7470A	PPB	25 Jan 2023 10:31:59	S	Std
2.5 - 1	52.5	10388	-	2.5000	7470A	PPB	25 Jan 2023 10:34:15	S	Std
5.0 - 1	55	19240	-	5.0000	7470A	PPB	25 Jan 2023 10:36:32	S	Std
7.5 - 1	57.5	28632	-	7.5000	7470A	PPB	25 Jan 2023 10:38:48	S	Std
10.0 - 1	510	39101	-	10.0000	7470A	PPB	25 Jan 2023 10:46:17	S	Std
ICV17 - 1	ICV17	15169	3.8980	-	7470A	PPB	25 Jan 2023 10:55:37	U	SMPL
ICB17 - 1	ICB17	-49	-0.0300	-	7470A	PPB	25 Jan 2023 10:57:52	U	SMPL
CCV18 - 1	CCV18	18657	4.7984	-	7470A	PPB	25 Jan 2023 11:00:07	U	SMPL
CCB18 - 1	CCB18	-65	-0.0341	-	7470A	PPB	25 Jan 2023 11:02:22	U	SMPL
CRA - 1	CRA	637	0.1471	-	7470A	PPB	25 Jan 2023 11:04:37	U	SMPL
HighStd - 1	HighStd	38031	9.7991	-	7470A	PPB	25 Jan 2023 11:06:52	U	SMPL
ChkStd - 1	ChkStd	26746	6.8863	-	7470A	PPB	25 Jan 2023 11:09:08	U	SMPL
PB150434BL - 1	PBW	-61	-0.0331	-	7470A	PPB	25 Jan 2023 11:11:24	U	SMPL
PB150434BS - 1	LCSW	13928	3.5777	-	7470A	PPB	25 Jan 2023 11:16:27	U	SMPL
O1260-02 - 1	TP-21	28	-0.0101	-	7470A	PPB	25 Jan 2023 11:18:44	U	SMPL
O1260-05 - 1	TP-22	45	-0.0057	-	7470A	PPB	25 Jan 2023 11:21:01	U	SMPL
O1260-08 - 1	TP-23	-5	-0.0186	-	7470A	PPB	25 Jan 2023 11:23:18	U	SMPL
O1260-11 - 1	TP-24	-22	-0.0230	-	7470A	PPB	25 Jan 2023 11:25:35	U	SMPL
O1260-14 - 1	TP-25	-17	-0.0217	-	7470A	PPB	25 Jan 2023 11:27:50	U	SMPL
CCV19 - 1	CCV19	19109	4.9150	-	7470A	PPB	25 Jan 2023 11:30:06	U	SMPL
CCB19 - 1	CCB19	-68	-0.0349	-	7470A	PPB	25 Jan 2023 11:32:21	U	SMPL
O1260-17 - 1	TP-26	13	-0.0140	-	7470A	PPB	25 Jan 2023 11:34:37	U	SMPL
O1260-20 - 1	TP-27	-5	-0.0186	-	7470A	PPB	25 Jan 2023 11:36:53	U	SMPL
O1260-23 - 1	TP-28	-1	-0.0176	-	7470A	PPB	25 Jan 2023 11:39:08	U	SMPL
O1266-02 - 1	VNJ-219	6	-0.0158	-	7470A	PPB	25 Jan 2023 11:41:24	U	SMPL
O1266-02DUP - 1	VNJ-219DUP	-14	-0.0209	-	7470A	PPB	25 Jan 2023 11:43:40	U	SMPL
O1266-02MS - 1	VNJ-219MS	13604	3.4941	-	7470A	PPB	25 Jan 2023 11:45:56	U	SMPL
O1266-02MSD - 1	VNJ-219MSD	13083	3.3596	-	7470A	PPB	25 Jan 2023 11:48:12	U	SMPL
PB150398TB - 1	PB150398TB	-80	-0.0380	-	7470A	PPB	25 Jan 2023 11:50:30	U	SMPL
O1266-02LX5 - 1		-8	-0.0194	-	7470A	PPB	25 Jan 2023 11:52:46	U	SMPL
O1266-02A - 1		16386	4.2122	-	7470A	PPB	25 Jan 2023 11:55:02	U	SMPL
CCV20 - 1	CCV20	19342	4.9752	-	7470A	PPB	25 Jan 2023 11:57:18	U	SMPL
CCB20 - 1	CCB20	-52	-0.0307	-	7470A	PPB	25 Jan 2023 11:59:33	U	SMPL