


LS123796

SFAM01.1 INSTRUMENT ID: CVL

Linear

μ Abs.:

38991



A= 0.0000e+000

B= 2.5672e-004 slope

C= -5.8016e-002 y-intercept

Rho= 0.9998859

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.016	0.016	287	0.000	287					
0.05	0.050					185					
0.20	0.200	0.200	-0.000	1004	0.0 %	1004	0				
2.50	2.500	2.489	-0.011	9921	0.0 %	9921					
5.00	5.000	4.935	-0.065	19450	0.0 %	19450					
7.50	7.500	7.609	0.109	29863	0.0 %	29863					
10.0	10.000	9.952	-0.048	38991	0.0 %	38991					

1001-0

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	287	-		0 SFAM01.1	PPB	1/25/2023 13:46	Std
	0.2 S01	1004	-		0.2 SFAM01.1	PPB	1/25/2023 13:48	Std
	2.5 S02	9921	-		2.5 SFAM01.1	PPB	1/25/2023 13:50	Std
	5 S03	19450	-		5 SFAM01.1	PPB	1/25/2023 13:53	Std
	7.5 S04	29863	-		7.5 SFAM01.1	PPB	1/25/2023 13:55	Std
	10 S05	38991	-		10 SFAM01.1	PPB	1/25/2023 13:57	Std
ICV018	ICV018	15267	3.8614 -		SFAM01.1	PPB	1/25/2023 14:00	SMPL
ICB018	ICB018	297	0.0182 -		SFAM01.1	PPB	1/25/2023 14:02	SMPL
CCV021	CCV021	19447	4.9345 -		SFAM01.1	PPB	1/25/2023 14:05	SMPL
CCB021	CCB021	263	0.0095 -		SFAM01.1	PPB	1/25/2023 14:07	SMPL
PB150461BL	PBW461	14	-0.0544 -		SFAM01.1	PPB	1/25/2023 14:09	SMPL
O1194-01	C0CP6	8	-0.056 -		SFAM01.1	PPB	1/25/2023 14:11	SMPL
O1194-02	C0CQ0	2	-0.0575 -		SFAM01.1	PPB	1/25/2023 14:14	SMPL
O1194-03	C0CQ3	8	-0.056 -		SFAM01.1	PPB	1/25/2023 14:16	SMPL
O1194-04	C0CQ5	13	-0.0547 -		SFAM01.1	PPB	1/25/2023 14:18	SMPL
O1194-05	C0CQ9	-10	-0.0606 -		SFAM01.1	PPB	1/25/2023 14:21	SMPL
O1194-06	C0CR1	-3	-0.0588 -		SFAM01.1	PPB	1/25/2023 14:23	SMPL
O1194-07	C0CR1D	-11	-0.0608 -		SFAM01.1	PPB	1/25/2023 14:25	SMPL
O1194-08	C0CR1S	4315	1.0497 -		SFAM01.1	PPB	1/25/2023 14:27	SMPL
O1194-09	C0CP2	10969	2.758 -		SFAM01.1	PPB	1/25/2023 14:32	SMPL
CCV022	CCV022	19071	4.838 -		SFAM01.1	PPB	1/25/2023 14:36	SMPL
CCB022	CCB022	253	0.0069 -		SFAM01.1	PPB	1/25/2023 14:39	SMPL
PB150462BL	PBW462	19	-0.0531 -		SFAM01.1	PPB	1/25/2023 14:41	SMPL
O1195-01	C0CP7	243	0.0044 -		SFAM01.1	PPB	1/25/2023 14:43	SMPL
O1195-02	C0CP9	270	0.0113 -		SFAM01.1	PPB	1/25/2023 14:45	SMPL
O1195-03	C0CQ1	287	0.0157 -		SFAM01.1	PPB	1/25/2023 14:48	SMPL
O1195-04	C0CQ4	305	0.0203 -		SFAM01.1	PPB	1/25/2023 14:50	SMPL
O1195-05	C0CQ6	300	0.019 -		SFAM01.1	PPB	1/25/2023 14:52	SMPL
O1195-06	C0CR0	299	0.0187 -		SFAM01.1	PPB	1/25/2023 14:55	SMPL
O1195-07	C0CR2	700	0.1217 -		SFAM01.1	PPB	1/25/2023 15:02	SMPL
O1195-08	C0CR2D	680	0.1166 -		SFAM01.1	PPB	1/25/2023 15:04	SMPL
O1195-09	C0CR2S	4793	1.1725 -		SFAM01.1	PPB	1/25/2023 15:11	SMPL
O1195-10	C0CP3	27392	6.9742 -		SFAM01.1	PPB	1/25/2023 15:18	SMPL
CCV023	CCV023	18723	4.7486 -		SFAM01.1	PPB	1/25/2023 15:20	SMPL
CCB023	CCB023	261	0.009 -		SFAM01.1	PPB	1/25/2023 15:22	SMPL