



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB123800
Analysis Date: 01/25/2023
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: Iwona
Extraction Date: 01/25/2023
Extraction IN Time: 13:50
Extraction OUT Time: 14:35
Thermometer ID: EXZT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB123800BL	LB123800BL	WATER	1.3	1000	100	3.5896	3.5896	3.02	3.5897	3.5897	0.0001	0.1
2	LB123800BS	LB123800BS	WATER	1.3	1000	100	3.1258	3.1258	3.01	3.1427	3.1427	0.0169	16.9
3	LB123800BSD	LB123800BSD	WATER	1.3	1000	100	3.5687	3.5687	3.03	3.5857	3.5857	0.0170	17
4	O1191-01	GRAB	WATER	1.6	1000	100	3.0174	3.0174	3.02	3.0519	3.0519	0.0345	34.5
5	O1290-01	GRAB	WATER	1.6	1000	100	3.0162	3.0162	3.04	3.0632	3.0632	0.0470	47



QC Batch# LB123800

Test: TPH

Analysis Date: 01/25/2023

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2969
pH Paper 0-14	M4909
Sodium Sulfate	EP2279
1:1 HCL	WP100035
Silica Gel	W2839
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP100036
LCSWD	5.00 ML	WP100037
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

Before Analysis

0.0020 gram Balance: 0.0018 (0.0018-0.0022) In OVEN TEMP1 : 70 °C Dessicator Time In1 : 15:41

1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time1: 15:00

Bal Check Time: 14:10 Out OVEN TEMP1: 70 °C Dessicator Time Out1: 16:25

Out Time1: 15:40

After Analysis

0.0020 gram Balance: 0.0021 (0.0018-0.0022) In OVEN TEMP2 : 71 °C Dessicator Time In2 : 17:31

1.0000 gram Balance: 1.0005 (0.9950-1.0050) In Time2: 17:00

Bal Check Time: 18:05 Out OVEN TEMP2: 71 °C Dessicator Time Out2: 18:00

Out Time2: 17:30

13123800

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tph 01290 WorkList ID : 166849 Department : Wet-Chemistry Date : 01-25-2023 13:32:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/31/2023	Water	O1191-01	TPH	Conc H2SO4 tr	ARAM01	N11	GRAB	01/18/2023	1664A
02/07/2023	Water	O1290-01	TPH	Conc H2SO4 tr	ARAM01	N11	GRAB	01/25/2023	1664A

Date/Time 01-25-23 13:35
Raw Sample Received by: JP (WOC)
Raw Sample Relinquished by: JTCM

Date/Time 01-25-23 16:30
Raw Sample Received by: JTCM
Raw Sample Relinquished by: JPCWC