

LB121811


7471B

INSTRUMENT ID: CV4

Linear

μ Abs.:

51772



A= 0.0000e+000

B= 1.9212e-004

C= 1.5108e-002

Rho= 0.9999539

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/11
0.0	0.000	0.016	0.016	7	0.000	7					
0.2	0.200	0.196	-0.004	942	0.0 %	942					12
2.5	2.500	2.469	-0.031	12771	0.0 %	12771					71
5.0	5.000	4.989	-0.011	25891	0.0 %	25891					0
7.5	7.500	7.568	0.068	39315	0.0 %	39315					0
10.0	10.000	9.961	-0.039	51772	0.0 %	51772					

LB123811 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Jan 2023 10:20:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	86	7	-	0.0000	7471B	PPB	26 Jan 2023 10:34:17	S	Std
0.2 - 1	86.2	942	-	0.2000	7471B	PPB	26 Jan 2023 10:36:34	S	Std
2.5 - 1	52.5	12771	-	2.5000	7471B	PPB	26 Jan 2023 10:38:51	S	Std
5.0 - 1	85	25891	-	5.0000	7471B	PPB	26 Jan 2023 10:41:08	S	Std
7.5 - 1	59.5	39315	-	7.5000	7471B	PPB	26 Jan 2023 10:43:24	S	Std
10.0 - 1	57	51772	-	10.0000	7471B	PPB	26 Jan 2023 10:45:40	S	Std
ICV19 - 1	ICV19	18935	3.6528	-	7471B	PPB	26 Jan 2023 10:48:31	U	SMPL
ICB19 - 1	ICB19	18	0.0186	-	7471B	PPB	26 Jan 2023 10:50:46	U	SMPL
CCV24 - 1	CCV24	25486	4.9114	-	7471B	PPB	26 Jan 2023 10:53:01	U	SMPL
CCB24 - 1	CCB24	-15	0.0122	-	7471B	PPB	26 Jan 2023 10:55:16	U	SMPL
CRA - 1	CRA	910	0.1899	-	7471B	PPB	26 Jan 2023 10:57:31	U	SMPL
HighStd - 1	HighStd	50229	9.6650	-	7471B	PPB	26 Jan 2023 10:59:46	U	SMPL
ChkStd - 1	ChkStd	36337	6.9961	-	7471B	PPB	26 Jan 2023 11:02:02	U	SMPL
PB150477BL - 1	PBS	-45	0.0065	-	7471B	PPB	26 Jan 2023 11:04:18	U	SMPL
PB150477BS - 1	LCSS	23451	4.5205	-	7471B	PPB	26 Jan 2023 11:06:34	U	SMPL
O1270-01 - 1	JC-1-012323	9179	1.7786	-	7471B	PPB	26 Jan 2023 11:08:50	U	SMPL
O1271-01 - 1	SU-3-012423	1191	0.2439	-	7471B	PPB	26 Jan 2023 11:11:06	U	SMPL
O1279-01 - 1	CL-1-012423	2090	0.4166	-	7471B	PPB	26 Jan 2023 11:13:23	U	SMPL
O1283-01 - 1	TP-H	3410	0.6702	-	7471B	PPB	26 Jan 2023 11:15:40	U	SMPL
O1283-04 - 1	MH-10	452	0.1019	-	7471B	PPB	26 Jan 2023 11:17:56	U	SMPL
CCV25 - 1	CCV25	23736	4.5752	-	7471B	PPB	26 Jan 2023 11:20:11	U	SMPL
CCB25 - 1	CCB25	-41	0.0072	-	7471B	PPB	26 Jan 2023 11:22:27	U	SMPL
O1285-01 - 1	B-07-SB01	14946	2.8865	-	7471B	PPB	26 Jan 2023 11:24:42	U	SMPL
O1285-02 - 1	B-07-SB02	33939	6.5354	-	7471B	PPB	26 Jan 2023 11:28:21	U	SMPL
O1289-01 - 1	VNJ209-STOCK	1440	0.2918	-	7471B	PPB	26 Jan 2023 11:30:35	U	SMPL
O1293-01 - 1	PL-3-012523	1498	0.3029	-	7471B	PPB	26 Jan 2023 11:32:51	U	SMPL
O1293-01DUP - 1	PL-3-012523DUP	1353	0.2750	-	7471B	PPB	26 Jan 2023 11:35:08	U	SMPL
O1293-01MS - 1	PL-3-012523MS	25368	4.8887	-	7471B	PPB	26 Jan 2023 11:37:24	U	SMPL
O1293-01MSD - 1	PL-3-012523MSD	25775	4.9669	-	7471B	PPB	26 Jan 2023 11:42:14	U	SMPL
O1310-01 - 1	S-1	1457	0.2950	-	7471B	PPB	26 Jan 2023 11:44:31	U	SMPL
O1293-01LX5 - 1		255	0.0641	-	7471B	PPB	26 Jan 2023 11:46:47	U	SMPL
O1293-01A - 1		22827	4.4006	-	7471B	PPB	26 Jan 2023 11:49:03	U	SMPL
CCV26 - 1	CCV26	25683	4.9493	-	7471B	PPB	26 Jan 2023 11:53:50	U	SMPL
CCB26 - 1	CCB26	-63	0.0030	-	7471B	PPB	26 Jan 2023 11:56:05	U	SMPL