

LB123818


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

47077



A= 0.0000e+000

B= 2.1244e-004 slope

C= 3.4746e-002 intercept

Rho= 0.9999675

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.00	0.000	0.038	0.038	14	0.000	14					
0.05	0.050					185	0				
0.20	0.200	0.208	0.008	817	0.0 %	817					
2.50	2.500	2.469	-0.031	11458	0.0 %	11458					
5.00	5.000	4.963	-0.037	23197	0.0 %	23197					
7.50	7.500	7.486	-0.014	35076	0.0 %	35076					
10.0	10.000	10.036	0.036	47077	0.0 %	47077					

0.0
1.4
1.00

LB123818

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	14	-		0 SFAM01.1	PPB	1/26/2023 15:15	Std
	0.2 S01	817	-		0.2 SFAM01.1	PPB	1/26/2023 15:17	Std
	2.5 S02	11458	-		2.5 SFAM01.1	PPB	1/26/2023 15:19	Std
	5 S03	23197	-		5 SFAM01.1	PPB	1/26/2023 15:21	Std
	7.5 S04	35076	-		7.5 SFAM01.1	PPB	1/26/2023 15:26	Std
	10 S05	47077	-		10 SFAM01.1	PPB	1/26/2023 15:28	Std
ICV021	ICV021	18040	3.8672 -		SFAM01.1	PPB	1/26/2023 15:31	SMPL
ICB021	ICB021	19	0.0388 -		SFAM01.1	PPB	1/26/2023 15:33	SMPL
CCV030	CCV030	23501	5.0274 -		SFAM01.1	PPB	1/26/2023 15:36	SMPL
CCB030	CCB030	-11	0.0324 -		SFAM01.1	PPB	1/26/2023 15:38	SMPL
O1296-01	MCOAL3	556	0.1529 -		SFAM01.1	PPB	1/26/2023 15:42	SMPL
O1296-02	MCOAL4	643	0.1713 -		SFAM01.1	PPB	1/26/2023 15:45	SMPL
O1296-03	MCOAL5	528	0.1469 -		SFAM01.1	PPB	1/26/2023 15:47	SMPL
O1296-04	MCOAL6	1096	0.2676 -		SFAM01.1	PPB	1/26/2023 15:49	SMPL
O1296-05	MCOAL7	1174	0.2842 -		SFAM01.1	PPB	1/26/2023 15:52	SMPL
O1296-06	MCOAL8	943	0.2351 -		SFAM01.1	PPB	1/26/2023 15:54	SMPL
O1296-07	MCOAL9	383	0.1161 -		SFAM01.1	PPB	1/26/2023 15:56	SMPL
O1296-08	MCOAM0	1091	0.2665 -		SFAM01.1	PPB	1/26/2023 15:58	SMPL
O1296-09	MCOAM1	779	0.2002 -		SFAM01.1	PPB	1/26/2023 16:01	SMPL
O1296-10	MCOAM2	895	0.2249 -		SFAM01.1	PPB	1/26/2023 16:03	SMPL
O1296-11	MCOAM3	579	0.1578 -		SFAM01.1	PPB	1/26/2023 16:05	SMPL
O1296-12	MCOAM4	671	0.1773 -		SFAM01.1	PPB	1/26/2023 16:07	SMPL
O1296-13	MCOAM5	2484	0.5625 -		SFAM01.1	PPB	1/26/2023 16:10	SMPL
O1296-14	MCOAM6	1302	0.3113 -		SFAM01.1	PPB	1/26/2023 16:12	SMPL
O1296-15	MCOAM7	1095	0.2674 -		SFAM01.1	PPB	1/26/2023 16:14	SMPL
O1296-16	MCOAM8	1384	0.3288 -		SFAM01.1	PPB	1/26/2023 16:16	SMPL
O1296-17	MCOAM9	2956	0.6627 -		SFAM01.1	PPB	1/26/2023 16:19	SMPL
O1296-18	MCOAN0	862	0.2179 -		SFAM01.1	PPB	1/26/2023 16:21	SMPL
O1296-19	MCOAN1	1260	0.3024 -		SFAM01.1	PPB	1/26/2023 16:23	SMPL
CCV031	CCV031	23522	5.0318 -		SFAM01.1	PPB	1/26/2023 16:26	SMPL
CCB031	CCB031	-25	0.0294 -		SFAM01.1	PPB	1/26/2023 16:28	SMPL
O1296-20	MCOAN2	1009	0.2491 -		SFAM01.1	PPB	1/26/2023 16:30	SMPL
O1296-21	MCOAN2D	1010	0.2493 -		SFAM01.1	PPB	1/26/2023 16:32	SMPL
O1296-22	MCOAN2S	14247	3.0614 -		SFAM01.1	PPB	1/26/2023 16:35	SMPL
PB150487BL	PBS487	32	0.0415 -		SFAM01.1	PPB	1/26/2023 16:37	SMPL
O1297-01	MCOAN3	1108	0.2701 -		SFAM01.1	PPB	1/26/2023 16:39	SMPL
O1297-02	MCOAN4	701	0.1837 -		SFAM01.1	PPB	1/26/2023 16:41	SMPL
O1297-03	MCOAN5	1526	0.3589 -		SFAM01.1	PPB	1/26/2023 16:44	SMPL
O1297-04	MCOAN6	1087	0.2657 -		SFAM01.1	PPB	1/26/2023 16:46	SMPL
O1297-05	MCOAN7	1302	0.3113 -		SFAM01.1	PPB	1/26/2023 16:48	SMPL
O1297-06	MCOAN9	1032	0.254 -		SFAM01.1	PPB	1/26/2023 16:50	SMPL
O1297-07	MCOAP0	966	0.24 -		SFAM01.1	PPB	1/26/2023 16:53	SMPL
O1297-08	MCOAP1	798	0.2043 -		SFAM01.1	PPB	1/26/2023 16:55	SMPL
O1297-09	MCOAP2	587	0.1594 -		SFAM01.1	PPB	1/26/2023 16:57	SMPL
O1297-10	MCOAP3	773	0.199 -		SFAM01.1	PPB	1/26/2023 17:00	SMPL
O1297-11	MCOAP4	660	0.175 -		SFAM01.1	PPB	1/26/2023 17:02	SMPL

LB123818

INSTRUMENT ID : CV1

O1297-12	MC0AP5	720	0.1877 -	SFAM01.1	PPB	1/26/2023 17:04 SMPL
O1297-13	MC0AP8	572	0.1563 -	SFAM01.1	PPB	1/26/2023 17:06 SMPL
O1297-14	MC0AP9	908	0.2276 -	SFAM01.1	PPB	1/26/2023 17:09 SMPL
O1297-15	MC0AQ0	694	0.1822 -	SFAM01.1	PPB	1/26/2023 17:11 SMPL
O1297-16	MC0AQ1	614	0.1652 -	SFAM01.1	PPB	1/26/2023 17:13 SMPL
CCV032	CCV032	23528	5.0331 -	SFAM01.1	PPB	1/26/2023 17:15 SMPL
CCB032	CCB032	-29	0.0286 -	SFAM01.1	PPB	1/26/2023 17:18 SMPL
O1297-17	MC0AQ2	668	0.1767 -	SFAM01.1	PPB	1/26/2023 17:20 SMPL
O1297-18	MC0AQ4	560	0.1537 -	SFAM01.1	PPB	1/26/2023 17:22 SMPL
O1297-19	MC0AQ6	1023	0.2521 -	SFAM01.1	PPB	1/26/2023 17:25 SMPL
O1297-20	MC0AQ6D	858	0.217 -	SFAM01.1	PPB	1/26/2023 17:27 SMPL
O1297-21	MC0AQ6S	13794	2.9652 -	SFAM01.1	PPB	1/26/2023 17:29 SMPL
PB150486BL	PBS486	-39	0.0265 -	SFAM01.1	PPB	1/26/2023 17:31 SMPL
CCV033	CCV033	23633	5.0554 -	SFAM01.1	PPB	1/26/2023 17:34 SMPL
CCB033	CCB033	-30	0.0284 -	SFAM01.1	PPB	1/26/2023 17:36 SMPL