



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB123819

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	01/26/2023	16:30
2	CAL2	1	Water	NA	NA	20.7	7.00	01/26/2023	16:31
3	CAL3	1	Water	NA	NA	20.7	10.03	01/26/2023	16:33
4	ICV	1	Water	NA	NA	20.7	6.99	01/26/2023	16:35
5	CCV1	1	Water	NA	NA	20.7	2.01	01/26/2023	16:38
6	O1284-01	1	Water	NA	NA	20.8	6.52	01/26/2023	16:40
7	O1284-02	1	Water	NA	NA	20.5	8.84	01/26/2023	16:42
8	O1320-01	1	Water	NA	NA	20.8	7.96	01/26/2023	16:45
9	O1321-01	1	Water	NA	NA	20.1	6.38	01/26/2023	16:47
10	O1321-02	1	Water	NA	NA	20.1	6.57	01/26/2023	16:49
11	O1327-01	1	Water	NA	NA	20.5	7.56	01/26/2023	16:50
12	O1327-01DUP	1	Water	NA	NA	20.6	7.57	01/26/2023	16:51
13	CCV2	1	Water	NA	NA	20.8	12.03	01/26/2023	17:00

123819

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph o1320 w WorkList ID : 166892 Department : Wet-Chemistry Date : 01-26-2023 16:23:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample		Collect Date	Method
						Storage	Location		
01/24/2023	Water	O1284-01	pH	Cool 4 deg C	PSEG03	D11	MW-1D-PURGEWATER-A	01/24/2023	9040C
01/24/2023	Water	O1284-02	pH	Cool 4 deg C	PSEG03	D11	MW-1D-PURGEWATER-B	01/24/2023	9040C
01/25/2023	Water	O1320-01	pH	Cool 4 deg C	PSEG03	D12	TOTE-COMP-0126	01/25/2023	9040C
01/26/2023	Water	O1321-01	pH	Cool 4 deg C	PSEG03	D12	33902	01/26/2023	9040C
01/26/2023	Water	O1321-02	pH	Cool 4 deg C	PSEG03	D12	33903	01/26/2023	9040C
01/26/2023	Water	O1327-01	pH	Cool 4 deg C	PSEG03	D12	N48965	01/26/2023	9040C

Date/Time 0126.23 16:25
Raw Sample Received by: JDC
Raw Sample Relinquished by: RM CWS

Date/Time 0126.23 18:00
Raw Sample Received by: RM CWS
Raw Sample Relinquished by: JDC