



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB123822

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2980
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W2888
buffer solution pH 7 yellow	W2887
Buffer Solution, PH2 (500ml)	W2806
pH 12.00 Buffer	W2940

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	01/26/2023	14:35
2	CAL2	1	Water	NA	NA	20.7	6.99	01/26/2023	14:36
3	CAL3	1	Water	NA	NA	20.7	10.03	01/26/2023	14:37
4	ICV	1	Water	NA	NA	20.7	6.98	01/26/2023	14:40
5	CCV1	1	Water	NA	NA	20.6	2.01	01/26/2023	14:45
6	O1289-02	1	Solid	20.02	20	20.2	7.94	01/26/2023	14:50
7	O1318-02	1	Solid	20.03	20	20.7	11.01	01/26/2023	15:00
8	O1318-05	1	Solid	20.04	20	20.9	10.22	01/26/2023	15:05
9	O1318-05DUP	1	Solid	20.03	20	20.8	10.23	01/26/2023	15:06
10	CCV2	1	Water	NA	NA	20.7	12.03	01/26/2023	15:10

123822

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY O1318 WorkList ID : 166884 Department : Wet-Chemistry Date : 01-26-2023 14:20:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
01/25/2023	Solid	O1289-02	Corrosivity	Cool 4 deg C	PSEG03	D11	VNJ209-STOCK	01/25/2023	9045D
01/25/2023	Solid	O1318-02	Corrosivity	Cool 4 deg C	RUTW01	D12	DR-1-012523	01/25/2023	9045D
01/25/2023	Solid	O1318-05	Corrosivity	Cool 4 deg C	RUTW01	D12	RJ-1-012523	01/25/2023	9045D

Date/Time 01.26.23 14:30
 Raw Sample Received by: JD (WC)
 Raw Sample Relinquished by: RM (WC)

Date/Time 01.26.23 16:30
 Raw Sample Received by: RM (WC)
 Raw Sample Relinquished by: JP (WC)